

# Evaluation of the Left Ventricular Ejection Fraction with Gated IQ-SPECT Myocardial Perfusion Imaging

Troels Joergensen and Susanne Haase Hansson

Department of Clinical Physiology and Nuclear Medicine, Naestved Hospital, Naestved, Denmark

The aim of this study was to evaluate the assessment of the left ventricular ejection fraction (LVEF) in patients by gated IQ-SPECT. **Methods:** Twenty-eight patients were examined using gated  $^{99m}\text{Tc}$ -sestamibi IQ-SPECT. Two different reconstruction datasets were created using the same projection data. The number of iterations, subsets, and gaussian filtering were based on 2 different recommendations from the manufacturer. For each dataset, end diastolic volume, end systolic volume, and LVEF were calculated using 4DMSPECT. A multigated planar equilibrium radionuclide ventriculography (MUGA) study with  $^{99m}\text{Tc}$ -labeled red blood cells was used as a reference for the LVEF. **Results:** The values of the different datasets were tested using the Bland–Altman analysis method. The calculated mean and 95% limits of agreement for the LVEF when dataset 1 and 2 were compared were  $-1.1\%$  and  $\pm 15\%$  points; when dataset 1 was compared with MUGA, the mean was calculated to  $-3.1\%$  points and  $\pm 17\%$  points for the 95% limits of agreement. When dataset 2 and MUGA were compared, the mean was  $-4.2\%$  and 95% limits of agreement of  $\pm 18\%$  points. **Conclusion:** Neither of the gated reconstructed datasets analyzed with 4DMSPECT was comparable to LVEF estimated by MUGA, with a tendency to overestimate LVEF. However, large random variations of the end diastolic volume, end systolic volume, and LVEF between the 2 gated reconstructed datasets were found. The reconstructed datasets were not interchangeable. Thus, these values should only be used with great caution when evaluating the functional state of the heart.

**Key Words:** ejection fraction; IQ-SPECT; gated blood pool; myocardial perfusion imaging

**J Nucl Med Technol 2015; 43:193–200**

DOI: 10.2967/jnmt.115.155382

As a diagnostic technique in coronary artery disease, electrocardiography (ECG)-gated SPECT (gSPECT) data acquisition is an established method. Reconstruction and analysis of the data provide information of left ventricular perfusion, wall motion, wall thickness, and quantification of parameters as left ventricular end diastolic and systolic

volumes (EDV and ESV, respectively) and ejection fraction (LVEF). The study is normally performed using dual- or triple-head  $\gamma$ -camera systems with low-energy all-purpose or low-energy high-resolution (LEHR) collimators installed and with a  $90^\circ$  or L detector configuration (1). To generate adequate imaging statistics, the overall acquisition time can be 20–30 min. Developments in hardware and software continue improving gSPECT image quality (2,3). In December 2008, Siemens Healthcare announced the Food and Drug Administration clearance of their newly developed hardware and software package called IQ-SPECT, with the purpose of shortening the acquisition time considerably. The Siemens IQ-SPECT system consists of SMARTZOOM collimators (magnifying collimators with a complex design), gantry movement control, and special reconstruction software. SMARTZOOM collimators center on the heart, collecting up to 4 times more counts than LEHR collimators. These collimators magnify the heart while still capturing counts from the entire field of view (4,5). IQ-SPECT's cardiocentric orbit is centered on the heart instead of the gantry's mechanical center, ensuring that the heart is always in the SMARTZOOM collimators' magnification area. Thus, the system is able to reduce acquisition time from approximately 20 min to approximately 5 min with the same patient dose (6,7).

In our department, the IQ-SPECT system has been used routinely for clinical gated and nongated myocardial perfusion imaging (MPI) studies since the installation in 2011. The use of the new system has led to suspicion that the reliability of the LVEF using IQ-SPECT may be questioned. The aim of this study was to evaluate the LVEF using IQ-SPECT with 2 different reconstruction settings (Siemens<sup>original</sup> and Siemens<sup>new</sup>) and to explore how these values relate to the LVEF using multigated planar equilibrium radionuclide angiography (MUGA). MUGA is a well-established method first reported in 1971 by Strauss et al. (8), and it was earlier shown to be a simple, reproducible, and highly accurate method for determination of LVEF (9,10).

## MATERIALS AND METHODS

### Patients

Twenty-eight patients (12 men, 16 women; mean age, 65 y; age range, 43–82 y) scheduled for routine MPI using  $^{99m}\text{Tc}$ -sestamibi IQ-gSPECT/CT were prospectively included. Exclusion criteria included pregnancy, arrhythmia, and patient inability to give written commitment. Seventeen patients were referred with suspected

Received Feb. 3, 2015; revision accepted May 5, 2015.

For correspondence contact: Troels Joergensen, Department of Clinical Physiology and Nuclear Medicine, Naestved Hospital, Ringstedgade 61, 4700 Naestved, Denmark.

E-mail: [troe@regionsjaelland.dk](mailto:troe@regionsjaelland.dk)

Published online Jun. 25, 2015.

COPYRIGHT © 2015 by the Society of Nuclear Medicine and Molecular Imaging, Inc.

coronary artery disease and 11 with known coronary artery disease. In 12 patients, the stress test was performed as a physiologic treadmill test, and 16 patients had a pharmacologic stress test with dipyridamol. None of the patients required longer acquisition time or higher dose, and Siemens quality control check was met (11). A 2-d SPECT/CT protocol was applied. gSPECT was performed only for the stress MPI. The stress test was performed according to the international guidelines (1). If the stress MPI was normal, the rest MPI was not performed. A MUGA study was performed in addition to the MPI  $3.4 \pm 0.5$  d (range, 3–6 d) later as a method of reference for the LVEF. All patients were informed orally and in writing. Written consent was given by all the patients. The study was approved by The Regional Committee on Health Research Ethics.

#### Method A: Gated Stress $^{99m}\text{Tc}$ -Sestamibi SPECT/CT Study

**Acquisition.** Gated stress MPI was performed using a Symbia T16 SPECT/CT scanner (Siemens AG) equipped with SMARTZOOM collimators. Images were acquired approximately  $115 \pm 20$  min (range, 73–234 min) after injection of an average  $650 \pm 50$  MBq (range, 600–705 MBq) of  $^{99m}\text{Tc}$ -sestamibi. Gated IQ-SPECT images were acquired over  $208^\circ$  cardiac-centric orbits with 17 views per detector of 14 s. The radius of the orbit was 280 mm. The total acquisition time for the study was only 6 min (including CT). Images were acquired with low-dose CT for attenuation correction for the nongated images. Attenuation correction for gated images was not supported by the manufacturer. Additional acquisition settings are given in Table 1.

**Reconstruction.** After data acquisition was completed, the study was transferred to a Siemens Syngo Processing workstation for reconstruction. The projection data were reviewed for motion, and motion correction was applied, if necessary. The manufacturer's original recommendation for the reconstruction was used: Siemens Flash3D iteration reconstruction algorithm, 15 iterations, 2 subsets, and 10-mm gaussian filtering (Siemens<sup>original</sup>). Processing steps included determination of myocardial axes and boundaries and masking of the myocardium.

The Siemens Flash3D technology is based on the maximum-likelihood reconstruction using ordered subset. It uses a 3-dimensional beam model for collimation in the iterative process, providing increased accuracy over earlier models (2-dimensional ordered-subset expectation maximization). Correctly modeling the colli-

mation beam enables the distribution of the activity over the slices to be more accurately reconstructed. The Flash3D has furthermore been modified to include SMARTZOOM collimator and gantry movement. The gaussian filtering is applied to the reconstructed images to reach the desired trade-off of resolution versus image noise.

**Data Analysis.** The reconstructed gated dataset was loaded into the third program 4DMSPECT (version 2010.1.0.56; University of Michigan Center). This program measures the EDV, ESV, and LVEF and is described elsewhere (12). Automatic processing was initially used for all software. Tracing of the ventricular walls was visually evaluated by an experienced operator, and if necessary the ventricular border surrounding the ventricle was modified and reprocessed.

Two experienced nuclear medicine physicians processed each dataset independently, beginning with the projection images and continuing through reconstruction and gated SPECT analysis. The average of the values was calculated and used for further analysis. The interobserver variability was 1%–3% points for all LVEFs (data not presented here).

#### Method B: Gated Stress $^{99m}\text{Tc}$ -Sestamibi SPECT/CT Study

The same projection data acquired in method A were used to create a new reconstruction for all patients. The manufacturer's new recommendation for the reconstruction was used: Siemens Flash3D iteration reconstruction algorithm, 12 iterations, 1 subset, and 10-mm gaussian filtering (Siemens<sup>new</sup>). Data analysis is as described in method A.

#### Method C: Multigated Blood-Pool Imaging

International guidelines for determination of the LVEF using planar equilibrium radionuclide angiography was followed (13).

**Acquisition.** The patients' red blood cells were labeled in vitro with  $740 \pm 45$  MBq (range, 680–810 MBq) of  $^{99m}\text{Tc}$  and reinjected in the patient. After the injection of the labeled red blood cells, the MUGA was performed in the left anterior oblique projection ( $30^\circ$ – $45^\circ$ ). The data were acquired using a  $64 \times 64$  matrix with a Symbia S  $\gamma$  camera (Siemens AG) with LEHR collimators. Sixteen frames per R-R interval were used, and the R-R interval tolerance window was set to 20%.

**Data Analysis.** LVEF was measured by the standard program supplied by the manufacturer (eSoft MI Apps VE50A; Siemens

**TABLE 1**  
Settings for Acquisition and Data Processing for the 3 Methods

Setting	Method		
	A	B	C
Radiopharmacy	$^{99m}\text{Tc}$ -sestamibi	$^{99m}\text{Tc}$ -sestamibi	$^{99m}\text{Tc}$ -UltraTag (Mallinckrodt Inc.)
Dose ( $\pm 2$ SDs)	$650 \pm 50$ MBq	$650 \pm 50$ MBq	$740 \pm 45$ MBq
Collimator	SMARTZOOM	SMARTZOOM	LEHR
Bins per cardiac cycle	8	8	16
Acquisition time	5 min	5 min	20 min
Matrix	$128 \times 128$	$128 \times 128$	$64 \times 64$
Pixel size	4.8 mm	4.8 mm	5.4 mm
Zoom	1	1	1.78
Camera position	2 detectors, $208^\circ$ , 17 views	2 detectors, $208^\circ$ , 17 views	Single-head $45^\circ$ left anterior oblique/ best septal separation between ventricles by adjustment
Processing software	4DMSPECT	4DMSPECT	Siemens esoft (MIApps)
Reconstruction algorithm	Iterative Flash3D (15i2s)	Iterative Flash3D (12i1s)	—
Filtering	Gaussian 10 mm	Gaussian 10 mm	Low-pass filtering

**TABLE 2**

Average Value, SD, and Range for LVEF, EDV, and ESV for All Patients and for Each of the Methods Applied

Parameter	Method		
	A	B	C
EDV (mL)			
Mean $\pm$ SD	87.0 $\pm$ 44.2	76.3 $\pm$ 46.3	—
Range	36–276	22–223	—
ESV (mL)			
Mean $\pm$ SD	33.3 $\pm$ 40.7	29.3 $\pm$ 40.0	—
Range	1.5–227	5–223	—
LVEF (%)			
Mean $\pm$ SD	67.5 $\pm$ 13.4	68.6 $\pm$ 11.4	64.4 $\pm$ 12.1
Range	18–83	19–82	18–83

EDV and ESV cannot be estimated in MUGA study.

Medical Solutions). Ventricular and background regions of interest were created semiautomatically by the operator, with support of the cine loop and phase image for an accurate definition of valvular planes. Butterworth 0.55, volume curve smoothing, and curve fitting were used.

### Statistics

Mean values and SDs were calculated for LVEF for each method and for EDV and ESV for methods A and B. Scatterplots were drawn, and linear regression analysis was performed by least-squares fitting. The coefficient of determination,  $R^2$ , and the Pearson correlation coefficient,  $r$ , were calculated ( $r = \sqrt{R^2}$ ).

The similarity of the methods was tested according to the method of Bland–Altman (14–19).

The mean of the differences, the 95% limits of agreement, and the confidence intervals (CIs) for the mean and the 95% limits of agreement were calculated.

The distribution of the differences was compared with a normal distribution using the Kolmogorov–Smirnov test. The differences among the LVEF results were shown in absolute LVEF units, called percentage points (% points) not by percentage of LVEF. The statistical analysis was done using Excel 2003 (Microsoft).

To help us in the interpretation of the method comparison, we predefined a medically accepted limit. LVEF is an important parameter for prediction of poor long-term prognosis, and the accuracy and reproducibility of the estimated LVEF are of great importance (20–22). The current guidelines for treating patients

with cardiotoxic chemotherapy states that chemotherapy should be considered discontinued if the patient presents again with a drop in LVEF of 10% points or more (23).

Therefore, if the new method for measuring the LVEF is unlikely to give readings for a subject who differs more than 10% points from those obtained using the old method, we would rely on the measurements made by the new method, as differences smaller than this would not be affected in the clinical interpretation of the result. On the other hand, differences of 10% points or more would not be satisfactory because an error of this magnitude could lead to a change in patient management. For the Bland–Altman plot, this means that  $\pm 1.96$  SDs is  $\leq 10\%$ , or actually that the upper/lower CI for the 95% limit of agreement is  $\leq 10\%$ .

### RESULTS

For the 28 patients included in the study, the mean of the LVEF was, respectively,  $68\% \pm 26\%$ ,  $69\% \pm 22\%$ , and  $64\% \pm 24\%$  for method A, method B, and method C. The range of the LVEF was approximately 20%–83% (Table 2). Most of our patients had normal LVEF, as only 1 was below 50%.

#### Comparison of Volumes

The results of the statistical analysis of EDV and ESV are summarized in Tables 2–4 and Figures 1 and 2.

When EDV and ESV for methods A and B were compared, the Pearson correlation coefficients were in both cases high,  $r = 0.99$ , and showed good correlation. From the Bland–Altman plots, Figures 1 and 2, we found that EDV has a systematic error of 11 mL and the systematic error of ESV was 4 mL. EDV and ESV were therefore estimated lower in method B than method A. The Bland–Altman analysis revealed high limits of agreement: 22 mL for EDV and 20 mL for ESV, shown in Figures 1 and 2.

#### Comparison of LVEF

On the Bland–Altman plots, Figures 3–5, we found that the highest systematic error for the LVEF was 4.2% points comparing method B with C, followed by 3.1% points comparing method A with C, and the smallest systematic error was 1.1% points comparing method A with B. In addition to the systematic error, the Bland–Altman plots indicated a significant random variance. For LVEF, the range of limits of agreement were unacceptably high

**TABLE 3**

Results of Kolmogorov–Smirnov Test for Each Method

Parameter	Method A vs. B, D(n) D(n, $\alpha$ )	Method A vs. C, D(n) D(n, $\alpha$ )	Method B vs. C, D(n) D(n, $\alpha$ )
LVEF, Kolmogorov–Smirnov test	0.05 < 0.24	0.09 < 0.24	0.10 < 0.24
EDV, Kolmogorov–Smirnov test	0.06 < 0.24	—	—
ESV, Kolmogorov–Smirnov test	0.08 < 0.24	—	—


If data are normally distributed, then critical value D(n, $\alpha$ ) will be larger than D(n). D(n, $\alpha$ ) is found in Kolmogorov–Smirnov table, for  $n = 28$  and  $\alpha = 0.05$  D(n, $\alpha$ ) = 0.24. D(n) is calculated for each method comparison and is in all cases lower than D(n, $\alpha$ ). All data are normally distributed.

**TABLE 4**

Results of Statistical Analysis for EDV and ESV for Comparison Between Method A and B

Analysis	Method	
	A vs. B, EDV (mL)	A vs. B, ESV (mL)
<b>Bland-Altman</b>		
Mean $\pm$ 2 SDs	10.8 $\pm$ 22	4.0 $\pm$ 20
CI of mean	$\pm$ 1.9	$\pm$ 1.9
CI of limits	$\pm$ 3.3	$\pm$ 3.2
<b>Linear regression</b> ( $y = ax + b$ )		
a	1.04	0.97
b	-14	-2.86
$R^2$	0.98	0.99
r	0.99	0.99

(respectively,  $\pm 15\%$  points,  $\pm 17\%$  points, and  $\pm 18\%$  points for method A vs. B, A vs. C, and B vs. C). There was a poor correlation between methods A and B for assessment of the LVEF ( $r = 0.71$ ) (Table 5), and LVEF values calculated from 4DMSPECT showed poor to modest correlation with MUGA ( $r = 0.86$  and  $0.67$ ).




**FIGURE 1.** Method A vs. B for EDV. (A) Scatterplot. Linear regression was calculated and is shown as full curve, and line of equality is illustrated by dotted curve. (B) Bland-Altman plot. One point in Bland-Altman plot is omitted due to scaling of x-axis. Bold curve = mean of differences; dotted curve =  $\pm 1.96$  SDs; small dotted curves = CIs of mean and CIs of limit of agreement.

## DISCUSSION

This study evaluated the estimation of LVEF using IQ-SPECT with different reconstruction settings.


MUGA was chosen as the method of reference, because it remains a generally accepted standard against which other LVEF measurement techniques are evaluated (24–26), and the method has been shown to be as good as MR imaging (27). Twenty-eight patients were included in the study. At this number, we estimated that the width of the 95% limit of agreement was so high that increasing the number of patients would only reduce the CIs and not contribute to reducing the fluctuations around the mean to the medically accepted tolerance.

When the IQ-SPECT system was introduced in 2011, recommendations from the manufacturer for processing of gated studies were to use Siemens Flash3D iterative reconstruction algorithm with 15 iterations and 2 subsets. This process resulted in reconstruction times of 10–12 min per study. To reduce reconstruction time, a new method was later suggested, reducing the number of iterations from 15 to 12 and the number of subsets from 2 to 1. To see the effect of



**FIGURE 2.** Method A vs. B for ESV. (A) Scatterplot. Linear regression was calculated and is shown as full curve, and line of equality is illustrated by dotted curve. (B) Bland-Altman plot. One point in Bland-Altman plot is omitted due to scaling of x-axis. Bold curve = mean of differences; dotted curve =  $\pm 1.96$  SDs; small dotted curves = CIs of mean and CIs of limit of agreement.





**FIGURE 3.** Method A vs. B for LVEF. (A) Scatterplot. Linear regression was calculated and is shown as full curve, and line of equality is illustrated by dotted curve. (B) Bland-Altman plot. Bold curve = mean of differences; dotted curve =  $\pm 1.96$  SDs; small dotted curves = CIs of mean and CIs of limit of agreement; gray area = predefined medical limits of 10% points.

changing reconstruction parameters, we compared method A with method B. This study reveals only a modest correlation ( $r = 0.83$ ,  $y = 0.7$ ,  $b = 20$ ) between the 2 methods.


Using the Bland-Altman analysis, we calculated the mean to  $-1.1\%$  points. Such a small systematic difference is within our predefined medical limits; however, the variations around the mean cause problems. As seen in the Bland-Altman plot in Figure 3, the differences between the 2 methods are widely spread, and the values for the  $\pm 1.96$  SDs are high, with values of  $\pm 15\%$  points. These variations around the mean tell us how far apart measurements by the 2 methods are likely to be.

Thus, for 95% of the patients the LVEF determined by method B will be within a range of  $+14\%$  point and  $-16\%$  points of the LVEF determined by method A.


As shown in Figure 3, the 95% limits of agreement are much wider than the medically accepted limits (the gray area). To accept the 2 methods to be interchangeable, the width of the 95% limits of agreement (and its lower/upper CIs) needs to be less than or equal to the predefined medical limit. In our case, the width of the 95% limits of agreement is 30% points (40% with the CI), which is 3 times (4 times with the CI) larger than our tolerance.

The statistical analysis of the EDV and ESV is summarized in Tables 2–4 and Figures 1 and 2. There is a large systematic error of 11 mL for the estimation of the EDV between the 2 methods and only a small systematic error of 4 mL for the ESV. The limits of agreement are in both cases approximately 20 mL, indicating large fluctuation around the mean. The large limit of agreement for the LVEF between method A and B is due to the large systematic error of the EDV and due to the variations in both EDV and ESV.

To demonstrate this variation, an example of reconstructed IQ-gSPECT data using methods A and B loaded into 4DMSPECT is displayed in Figures 6 and 7. Method A has better/more clearly defined borders than method B, and the volume of the myocardium seems to be smaller. As the number of iterations and subsets increase, the level of details in the image (including edge sharpness and conspicuity) is expected to improve but also the noise is expected to increase. When the same gaussian filtering is used, the images in method B are in this case oversmoothed (reducing the gaussian filtering has little effect). In Figure 7, the 4DMSPECT seems to have a problem, particularly in method B, with the definition of the borders used for volume estimation.



**FIGURE 4.** Method A vs. C for LVEF. (A) Scatterplot. Linear regression was calculated and is shown as full curve, and line of equality is illustrated by dotted curve. (B) Bland-Altman plot. Bold curve = mean of differences; dotted curve =  $\pm 1.96$  SDs; small dotted curves = CIs of mean and CIs of limit of agreement; gray area = predefined medical limits of 10% points.



**FIGURE 5.** Method B vs. C for LVEF. (A) Scatterplot. Linear regression was calculated and is shown as full curve, and line of equality is illustrated by dotted curve. (B) Bland-Altman plot. Bold curve = mean of differences; dotted curve =  $\pm 1.96$  SDs; small dotted curves = CIs of mean and CIs of limit of agreement; gray area = predefined medical limits of 10% points.

Overall, the statistical analysis and example show that with even the most optimistic interpretation there are considerable discrepancies between the 2 methods, and we think that the disagreement is unacceptable for clinical use.

Another question to be answered is whether the LVEF measured by either method A or method B is comparable to the method of reference. When method A was compared

with method C, the scatterplot in Figure 4 and Table 5 shows that the correlation between the 2 methods was poor. The Pearson correlation coefficient is only 0.77, with a y intercept at 12 and a gradient of 0.86. The deviation from the identity line is obvious. From the Bland-Altman analysis, the mean difference is calculated to  $-3.1\%$  points, which are within our predefined limit. As before, the variations around the mean are high, and the  $\pm 1.96$  SDs are equal to  $\pm 17\%$  points. In this case, we can estimate that for 95% of the patients examined, the LVEF determined by method A will be between 14% points above the method of reference and 20% points below. The limit of our medically accepted error of 10% points is thus exceeded by a factor of more than 3, implying that method A cannot be used as an alternative for estimation of the LVEF, compared with method C.


Comparing method B and C is even worse. The Pearson correlation coefficient is as low as 0.70, and visual inspection of the scatterplot in Figure 5 confirms that the correlation is poor. The mean difference is calculated to  $-4.2\%$  points, and the limits of agreement are  $\pm 1.96$  SDs = 18.2% points. Once again, we estimated that for 95% of the subjects the LVEF determined by method B would be between 14% points above method of reference and 22% points below. The predefined medical limit was exceeded with a factor close to 4, and we must conclude that methods B and C are not interchangeable.

In summary, none of the values EDV, ESV, and LVEF generated by the methods A, B, and C are comparable because of the large variations.

To the best of our knowledge, most of the studies performed with IQ-SPECT are nongated SPECT MPI, comparing image quality with conventional SPECT MPI (28–31). Onishi et al. (31) showed that the spatial resolution in the center of the scanner and image quality of the IQ-SPECT was comparable to the conventional SPECT (in a radius of 28 cm), suggesting that IQ-SPECT would be the optimal technology for MPI because of the reduced acquisition time. However, IQ-gSPECT's ability to quantify the EDV, ESV, and LVEF was not investigated because gated phantom studies were not performed. Corbette et al. (28) found in a single-center clinical trial that IQ-SPECT

**TABLE 5**  
Results of Statistical Analysis for LVEF for Each Comparison Between Methods A, B, and C

Analysis	Methods		
	A vs. B (% points)	A vs. C (% points)	B vs. C (% points)
Bland-Altman			
Mean $\pm 2$ SDs	$-1.1 \pm 15.0$	$-3.1 \pm 17.2$	$-4.2 \pm 18.2$
CI of mean	$\pm 2.8$	$\pm 3.3$	$\pm 3.4$
CI of limits	$\pm 4.9$	$\pm 5.6$	$\pm 5.9$
Linear regression ( $y = ax + b$ )			
a	0.71	0.86	0.67
b	20.6	12.2	25.6
$R^2$	0.69	0.60	0.50
r	0.83	0.77	0.70




**FIGURE 6.** Screen capture of tap setup in 4DMSPECT. Determination of position of left ventricular (LV) center and apical and basal limits. (A) Results for method A. (B) Results for method B. Quality of images in method B is in this case reduced. Gaussian filtering in method A and B is equal.

provided better image quality than conventional SPECT, but again gated studies were not performed. Caobelli et al. (29,30) concluded that MPI with IQ-SPECT protocol can be acquired using about a quarter the scan time normally needed without disagreement, compared with full-time scan acquisition performed with standard protocols but only for nongated studies. A paper from Siemens Healthcare (Hawman et al. (7)) evaluated several patient studies to describe the differences between conventional SPECT and IQ-SPECT, but none of the studies was done as gSPECT, and estimation of EDV, ESV, and LVEF was lacking. An earlier paper from Siemens (11) compared the IQ-gSPECT with LEHR conventional gSPECT for estimation of the LVEF analyzed with 4DMSPECT. We have not asked Siemens Healthcare for permission to transmit the results. Talleruphuus et al. (32) have also compared the quantification of the EDV, ESV, and LVEF between IQ-gSPECT and conventional LEHR gSPECT. For conventional gSPECT the values for LVEF, ESD, and EDV were  $60.8 \pm 3.0\%$  points,  $44.2 \pm 6.6\%$  points, and  $101.6 \pm 10.1\%$  points. For IQ-gSPECT, the corresponding values were  $66.0 \pm 4.2\%$  points,  $32.2 \pm 6.2\%$  points, and  $79.2 \pm 9.3\%$  points. They concluded that IQ-gSPECT studies exhibited systematic deviations from

conventional studies concerning EDV, ESV, and LVEF (as estimated by Quantitative Gated SPECT software [QGS]).

A follow-up of our study would be to compare the LVEF estimated with other software packages (e.g., QGS, Emory toolbox) and try to optimize the reconstruction settings to see if it is possible to obtain values for the LVEF that are comparable to the LVEF estimated by method C.

Some limitations of this study should be noted. First, most of our patients had normal LVEF. The correlation in patients with impaired left-ventricular function should be further evaluated in the 15%–50% range. Second, the acquisition of gSPECT was performed more than 1 h after the pharmacologic stress test. The effect of myocardial stunning in the state of poststress was unknown. Third, there are known variations in the estimation of EDV, ESV, and LVEF between third-party cardiac software packages, caused by differences in the way the contours that identify the cardiac surfaces are generated (33). Only the software package 4DMSPECT has been used here. Fourth, quantification of LVEF is dependent on the number of gates per cardiac cycle. In methods A and B, 8 gates per cardiac cycle are used, and in method C 16 gates per cardiac cycle are used. In 1995, Germano et al. (34) showed that the use of 8 gates



**FIGURE 7.** Screen capture of IQ-gSPECT data loaded into 4DMSPECT. Frame 4 of 8 is shown. (A) Results displayed for method A, EDV = 58 mL, ESV = 10 mL, and LVEF = 83%. (B) Results displayed for method B, EDV = 42 mL, ESV = 14 mL, and LVEF = 67%. LVEF for method C = 73%.

per cardiac cycle instead of 16 resulted in a constant and predictable 4% points decrease in LVEF. However, this should not influence our results. The main problem is the LVEF variations between the 2 methods.

## CONCLUSION

Some differences in the measurement of LVEF can be expected when using different imaging techniques. To evaluate whether 2 methods were interchangeable, we defined an acceptable clinical change for the total deviation of LVEF, which was set to 10% points. With this limit in mind, we can conclude that changing the reconstruction settings for the iterative algorithm has a large impact on the estimation of EDV, ESV, and LVEF using IQ-gSPECT. Therefore, methods A and B are not interchangeable. Furthermore, we can conclude that neither LVEF value using method A nor that using method B (estimated with 4DMSPECT) is comparable with LVEF estimated by MUGA, and the methods are not suitable for evaluation of LVEF in critical settings, for example, in control of chemotherapy or evaluation of cardiac pumping efficiency.

## DISCLOSURE

No potential conflict of interest relevant to this article was reported.

## REFERENCES

- Hesse B, Tagel K, Cuocolo A, et al. EANM/ESC procedural guidelines for myocardial perfusion imaging in nuclear cardiology. *Eur J Nucl Med Mol Imaging*. 2005;32:855–897.
- Slomka PJ, Patton JA, Berman DS, Germano G. Advances in technical aspects of myocardial perfusion SPECT imaging. *J Nucl Cardiol*. 2009;16:255–276.
- DePuey EG. Advances in SPECT camera software and hardware: currently available and new on the horizon. *J Nucl Cardiol*. 2012;19:551–581.
- Zeintl J, Rempel TD, Bhattacharya M, Malmin RE, Vija AH. Performance characteristics of the SMARTZOOM collimator. *Nucl Sci Symp Med Imaging Conf*. 2011;2426–2429.
- Rajaram R, Bhattacharya M, Xinhong D, Malmin R. Tomographic performance characteristics of the IQSPECT system. *Nucl Sci Symp Med Imaging Conf*. 2011;2451–2456.
- Vija AH, Malmin R, Yahil A, et al. A method for improving the efficiency of myocardial perfusion imaging using conventional SPECT and SPECT/CT imaging systems. *Nucl Sci Symp Med Imaging Conf*. 2010;3433–3437.
- Hawman P, Ghosh P. *IQ-SPECT: A Technical and Clinical Overview*. White paper. Knoxville, TN: Siemens Medical Solutions, USA Molecular Imaging; 2012.
- Strauss HW, Zaret BL, Hurley BJ, Natarajan TK, Pitt B. A scintiphotographic method for measuring left ventricular ejection fraction in man without cardiac catheterization. *Am J Cardiol*. 1971;28:575–580.
- Hains AD, Khawaja IA, Hinge DA, Lahiri A, Raftery EB. Radionuclide left ventricular ejection fraction: a comparison of three methods. *Br Heart J*. 1987;57:242–246.
- Wackers FJT, Berger HJ, Johnstone DE, et al. Multiple gated cardiac blood pool imaging for left ventricular ejection fraction: validation of the technique and assessment of variability. *Am J Cardiol*. 1979;43:1159–1166.
- Siemens Healthcare. *IQ-SPECT Hints and Tips. Software Version VA60A*. Erlangen, Germany: Siemens HealthCare Headquarters; 2011.
- Ficaro EP, Lee BC, Kritzman JN, Corbett JR. Corridor4DM: the Michigan method for quantitative nuclear cardiology. *J Nucl Cardiol*. 2007;14:455–465.
- Hesse B, Lindhardt TB, Acampa W, et al. EANM/ESC guidelines for radionuclide imaging of cardiac function. *Eur J Nucl Med Mol Imaging*. 2008;35:851–885.
- Bland JM, Altman DG. Measuring agreement in method comparison studies. *Stat Methods Med Res*. 1999;8:135–160.
- Bland JM, Altman DG. Measurement in medicine: the analysis of method comparison studies. *Statistician*. 1983;32:307–317.
- Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet*. 1986;8476:307–310.
- Bland JM, Altman DG. Comparing methods of measurement: why plotting difference against standard method is misleading. *Lancet*. 1995;346:1085–1087.
- Dewitte K, Fierens C, Stöckl D, Thienpont LM. Application of the Bland-Altman plot for the interpretation of method-comparison studies: a critical investigation of its practice. *Clin Chem*. 2002;48:799–801.
- Stöckl D, Rodríguez D, Cabaleiro, Uytendaele K, Thienpont LM. Interpreting method comparison studies by use of the Bland-Altman plot: reflecting the importance of sample size by incorporating confidence limits and predefined error limits in the graphic. *Clin Chem*. 2004;50:2216–2218.
- Hammermeister KE, de Rouen TA, Dodge HT. Variables predictive of survival in patients with coronary disease: selection by univariate and multivariate analyses from the clinical, electrocardiographic, exercise, arteriographic, and quantitative angiographic evaluations. *Circulation*. 1979;59:421–430.
- White HD, Norris RM, Brown MA, Brandt PW, Whitlock M, Wild CJ. Left ventricular end-systolic volume as the major determinant of survival after recovery from myocardial infarction. *Circulation*. 1987;76:44–51.
- Yamaguchi A, Ino T, Adachi H, Murata S, Kamio H. Left ventricular volume predicts postoperative course in patient with ischemic cardiomyopathy. *Ann Thorac Surg*. 1998;65:434–438.
- Schwartz RG, McKenzie WB, Alexander J, et al. Congestive heart failure and left ventricular dysfunction complicating doxorubicin therapy: seven-year experience using serial radionuclide angiography. *Am J Med*. 1987;82:1109–1118.
- Hacker M, Hoyer X, Kupzyk S, La Fougere C, Kois J, Stempfle HU. Clinical validation of the gated blood pool SPECT QBS processing software in congestive heart failure patients: correlation with MUGA, first-pass RNV, and 2D-echocardiography. *Int J Cardiovasc Imaging*. 2006;22:407–416.
- De Bondt P, Winter OD, Sutter JD, Dierckx RA. Agreement between four available algorithms to evaluate global systolic left and right ventricular function from tomographic radionuclide ventriculography and comparison with planar imaging. *Nucl Med Commun*. 2005;26:351–359.
- Mondelli JA, Di Luzio S, Nagaraj A, Kane BJ, Smulevitz B, Nagaraj AV. The validation of volumetric real-time 3-dimensional echocardiography for the determination of left ventricular function. *J Am Soc Echocardiogr*. 2001;14:994–1000.
- San Román JA, Candell-Riera J, Arnold R, et al. Quantitative analysis of left ventricular function as a tool in clinical research: theoretical basis and methodology [article in Spanish]. *Rev Esp Cardiol*. 2009;62:535–551.
- Corbett J, Meden J, Ficaro E. Clinical validation of attenuation corrected cardiac imaging with IQSPECT SPECT/CT [abstract]. *J Nucl Med*. 2010;51(suppl 2):1722.
- Caobelli F, Pizzocaro C. Evaluation of patients with coronary artery disease. IQ-SPECT protocol in myocardial perfusion imaging: preliminary results. *Nuklearmedizin*. 2013;52:178–185.
- Caobelli F, Pizzocaro C, Massetti V, et al. Possible use of IQSPECT protocol in myocardial perfusion imaging in the evaluation of patients with suspected or known coronary artery disease: preliminary results. Paper presented at: the Annual Congress of the European Association of Nuclear Medicine; Lyon, France; October 19–23, 2013.
- Onishi H, Matsutomo N, Kiguchi Y, Sato T, Amijima H. Differential impact of multifocus fan beam collimation with L-mode and conventional systems on the accuracy of myocardial perfusion imaging: quantitative evaluation using phantoms. *Asia Oceania J Nucl Med Biol*. 2013;1:28–34.
- Talleruphuus U, Davidsdottir ML, Lonsdale MN. How smart is SMARTZOOM? A direct comparison between IQ-SPECT and traditional clinical myocardial perfusion SPECT. Paper presented at: the Annual Congress of the European Association of Nuclear Medicine; Lyon, France; October 19–23, 2013.
- Wolak A, Slomka PJ, Fish MB, et al. Quantitative myocardial-perfusion SPECT: comparison of three state-of-the-art software packages. *J Nucl Cardiol*. 2008;15:27–34.
- Germano G, Kiat H, Kavanagh PB, et al. Automatic quantification of ejection fraction from gated myocardial perfusion SPECT. *J Nucl Med*. 1995;36:2138–2147.





## Evaluation of the Left Ventricular Ejection Fraction with Gated IQ-SPECT Myocardial Perfusion Imaging

Troels Joergensen and Susanne Haase Hansson

*J. Nucl. Med. Technol.* 2015;43:193-200.

Published online: June 25, 2015.

Doi: 10.2967/jnmt.115.155382

---

This article and updated information are available at:  
<http://tech.snmjournals.org/content/43/3/193>

---

Information about reproducing figures, tables, or other portions of this article can be found online at:  
<http://tech.snmjournals.org/site/misc/permission.xhtml>

Information about subscriptions to JNMT can be found at:  
<http://tech.snmjournals.org/site/subscriptions/online.xhtml>

*Journal of Nuclear Medicine Technology* is published quarterly.  
SNMMI | Society of Nuclear Medicine and Molecular Imaging  
1850 Samuel Morse Drive, Reston, VA 20190.  
(Print ISSN: 0091-4916, Online ISSN: 1535-5675)

© Copyright 2015 SNMMI; all rights reserved.