

A New Tool for Molecular Imaging: The Microvolumetric β Blood Counter

Laurence Convert¹, Guillaume Morin-Brassard², Jules Cadorette¹, Mélanie Archambault¹, M'hamed Bentourkia¹, and Roger Lecomte¹

¹Department of Nuclear Medicine and Radiobiology, Université de Sherbrooke, Sherbrooke, Québec, Canada; and ²Department of Electrical and Computer Engineering, Université de Sherbrooke, Sherbrooke, Québec, Canada

Radiotracer kinetic modeling in small animals with PET allows absolute quantification of physiologic and biochemical processes in vivo. It requires blood and tissue tracer concentrations as a function of time. Manual sampling, the reference method for blood tracer concentration measurements, requires fairly large amounts of blood besides being technically difficult and time-consuming. An automated microvolumetric β blood counter (μ BC) was designed to circumvent these limitations by measuring the blood activity in real time with PET scanning. **Methods:** The μ BC uses direct β -particle detection to reduce its footprint and is entirely remote controlled for sampling protocol selection and real-time monitoring of measured parameters. Sensitivity has been determined for the most popular PET radioisotopes (¹⁸F, ¹³N, ¹¹C, ⁶⁴Cu). Dispersion within the sampling catheter has been modeled to enable automatic correction. Blood curves obtained with the μ BC were compared with manual samples and PET-derived data. The μ BC was used to estimate the myocardial blood flow (MBF) of mice injected with ¹³N-ammonia and to compare the myocardial metabolic rate of glucose (MMRG) of rats injected with ¹⁸F-FDG for arterial and venous cannulation sites. **Results:** The sensitivity limit ranges from 3 to 104 Bq/ μ L, depending on the isotope and the catheter used, and was found to be adequate for most small-animal studies. Automatic dispersion correction appears to be a good approximation of dispersion-free reference curves. Blood curves sampled with the μ BC are well correlated with curves obtained from manual samples and PET images. With correction for dispersion, the MBF of anesthetized mice at rest was found to be 4.84 ± 0.5 mL/g/min, which is comparable to values found in the literature for rats. MMRG values derived from the venous blood tracer concentration are underestimated by 60% as compared with those derived from arterial blood. **Conclusion:** The μ BC is a compact automated counter allowing real-time measurement of blood radioactivity for pharmacokinetic studies in animals as small as mice. Reliable and reproducible, the device makes it possible to increase the throughput of pharmacokinetic studies with reduced blood sample handling and staff exposure, contributing to speed up new drug development and evaluation.

Key Words: β -detector; blood sampling; small-animal PET; pharmacokinetic modeling

J Nucl Med 2007; 48:1197–1206
DOI: 10.2967/jnumed.107.042606

Nowadays, experimental studies in animals have become an essential part of research for new drug development and therapy assessment. Rats and mice are researchers' favorite animal models because their genome, which is very similar to the human one, has been entirely decoded and can be used for genetic manipulations (1). The mouse counted for >90% of the mammals used in research over the world in 2000 (2). Concurrently, most clinical imaging modalities have been adapted to the size of rats and mice, and the spatial resolution and sensitivity of dedicated PET scanners are now suitable to investigate biochemical and physiologic processes in vivo in rodents, including various metabolisms, receptor binding, tumor size and grade, or gene expression (2).

To extract quantitative biologic information from PET studies, the tissue and blood tracer concentration must be measured as a function of time (3,4). Whereas the absolute tracer concentration in tissue can be derived from PET images, several different methods must be used to obtain the blood time–activity curves (5). When a suitable blood pool can be identified within the field of view of the scanner, it can be derived from the sequence of high-resolution PET images (6,7). Otherwise, an external means must be used to measure the blood time–activity curve. Manual sampling, considered as the gold standard, is rather challenging in small animals because of the very small amounts of blood available and the required recovery time for repeated studies (2 wk for 1.6 mL in rats, 0.2 mL in mice) (8). Mini-PET systems over the wrist have been proposed for human studies, but the size and resolution of these devices is not adapted to small animals (9). More suitable external monitors for rodents were developed to address the problem (10–13), but none of these devices provides an integrated solution for real-time measurement of blood radiotracer concentration that is fully adapted to rats and mice.

Received Nov. 15, 2006; revision accepted Apr. 17, 2007.

For correspondence or reprints contact: Roger Lecomte, PhD, Department of Nuclear Medicine and Radiobiology, Université de Sherbrooke, 3001 12th Ave. N., CRC/CHUS #1864, Sherbrooke, Québec, Canada J1H 5N4.

E-mail: Roger.Lecomte@USherbrooke.ca

COPYRIGHT © 2007 by the Society of Nuclear Medicine, Inc.

In the present study, we describe a new online counting system for real-time blood time–activity curve measurement. On the basis of the direct detection of β -particles, its detection volume and overall footprint were expressly tailored for small-animal pharmacokinetic studies. Fully integrated with complete signal-processing electronics and software for real-time control, calibration, and dispersion correction, the device is also very sensitive to low radioactivity concentrations. The device was used for the measurement of myocardial blood flow (MBF) in mice and for the assessment of myocardial metabolic rate of glucose (MMRG) in rats with venous and arterial sampling sites.

MATERIALS AND METHODS

System Overview

The proposed microvolumetric blood counter (μ BC) is based on the direct detection of positrons and consists of 2 separated hardware units and control software (14). It can be used as stand-alone or hooked to the LabPET small-animal PET scanner (Gamma Medica-Ideas Inc.) (Fig. 1A).

The detection unit includes the β -detector made of 2 *p-i-n* photodiodes having a 3×30 mm² active area with signal-processing electronics. A thin tungsten shielding (<1.3 cm thick) divided in 2 interlocking parts with internal copper lining is provided to avoid electromagnetic interference (EMI) pickup and protect the photodiodes from radioactivity in the animal and injection syringe (Fig. 1B). To avoid stress, the lid where the catheter from the animal is placed remains motionless, whereas the base can be moved on a rail to close the assembly (Fig. 1B). Two inner linings are available to accommodate PE50 (0.58-mm inner diameter [ID], 0.965-mm outer diameter [OD]) and PE10 (0.28-mm ID, 0.61-mm OD) catheters (Intramedic polyethylene tubing; Becton Dickinson), generally used with rats and mice, respectively.

The main unit includes a pumping system and the control electronics for the pump, the detector unit, and the communication with a personal computer or the LabPET. The syringe pump (Microliter Flow M.P.C.; Harvard Apparatus) can accommodate speeds from 3×10^{-3} μ L/h (for a 1- μ L syringe) to 1.5 mL/min (for a 3-mL syringe).

The user-friendly control software, running under Windows (Microsoft), Linux, or MacOS (Apple) environments, was devel-

oped to set up experiment parameters, display the blood time–activity curve in real time, and process recorded data (Fig. 2). During an experiment, the user can choose to display the measured blood activity curve in nCi/ μ L, Bq/ μ L, or cps/ μ L (cps = counts per second). When the μ BC is hooked to the LabPET scanner, the counter control software is integrated to the scanner's user interface. Communication between the computer and the μ BC is made through any RS232 serial port or USB (universal serial bus) adapter. For a more complete technical description of the μ BC, refer to (14,15).

Data Corrections

To be used as the input function in pharmacokinetic models, the data recorded by the μ BC must be corrected for detector efficiency, radionuclide decay time, radioactivity dispersion inside the catheter, system dead time, and propagation delay between the cannulation point and the detector. Only detector efficiency and radioactivity dispersion issues will be addressed in this article, as other corrections and the physical performance evaluation of the device have been reported previously (14,15).

Sensitivity. Due to attenuation within blood, catheter wall, and photodiode dead layer, only a small fraction of the β -particles actually reaches the detector and is measured by the μ BC. The true count rate $C_T(t)$ in becquerels, reflecting the amount of radioactivity in the measured sample, can be computed from the measured count rate $C_m(t)$ in cps by:

$$C_T(t) = C_m(t)/K, \quad \text{Eq. 1}$$

where K is the sensitivity calibration factor in cps/Bq. Because detection sensitivity is a function of the radioisotope $E_{\beta\text{max}}$ and discriminator threshold DI , K must be evaluated for each radionuclide and computed for each experimental setup.

A set of experiments was conducted to determine the relationship between K and DI . A 4-cm-long sample of a 5.5-MBq/mL solution of ^{18}F -FDG was injected in a small piece of PE50 catheter and centered into the μ BC. The count rate was recorded as a function of DI and the sensitivity K computed. The resulting curve could be fitted by a simple linear relationship between a minimum threshold level, DI_{min} , set just above the electronic noise and a maximum threshold level, DI_{max} , set to keep μ BC sensitivity high enough for animal studies:




FIGURE 1. (A) μ BC hooked to LabPET scanner, together with injection pump on left. (B) Detection unit shows shielding and rail systems.




FIGURE 2. Control software, showing the acquisition (top) and configuration (bottom) tabs.

$$K(DI) = K_0 + s \times DI, \quad \text{Eq. 2}$$

where DI is expressed as a fraction of the maximum signal amplitude corresponding to $E_{\beta\text{max}}$, K_0 (cps/Bq) is the projected detector sensitivity for the discriminator level set to 0, and s is the slope. It was implemented in the control software as an 8-point curve defined between DI_{min} and DI_{max} with $K_{\text{opt}} = K(DI_{\text{min}})$. This protocol was repeated to obtain the optimum absolute sensitivity, K_{opt} (cps/Bq), with the 4 most commonly used PET isotopes in our center (^{18}F , ^{13}N , ^{11}C , and ^{64}Cu) for PE50 and PE10 catheters. As the typical background count rate is negligible in most animal studies (14), these experiments were performed assuming a default background count rate arbitrarily set to 5 cps.

The minimum detectable activity was then computed following 2 mathematic models. The first one, based on the Curie equation (16), determines the activity A_{Imin} (Bq/μL) that is statistically significant $\pm 5\%$ over the background count rate:

$$A_{\text{Imin}} = (4.653 \times \sqrt{n} + 2.706) / (K_{\text{opt}} \times V), \quad \text{Eq. 3}$$

where n (cps) is the background noise and V is the detection volume (7.926 μL for PE50 and 1.847 μL for PE10). The second model is less stringent but also common in radiation detection science. The minimum activity $A_{2\text{min}}$ (Bq/μL) is expressed as a count rate that is 3 times higher than the statistical noise:

$$A_{2\text{min}} = (3 \times \sqrt{n}) / (K_{\text{opt}} \times V). \quad \text{Eq. 4}$$

The maximum activity that can be detected by the μBC was also computed as follows:

$$A_{\text{max}} = N_{\text{max}} / (K_{\text{opt}} \times V), \quad \text{Eq. 5}$$

where $N_{\text{max}} = 65,535$ cps is the maximum count rate permitted by the μBC 16-bit counter.

Dispersion. Dispersion effects need to be considered when radioactivity is moving inside catheters. Indeed, a fast sharp radioactivity edge, as encountered with bolus injections, can be highly smoothed as it travels inside the catheter and because of its

transit time in the μBC , resulting in severe distortions of estimated metabolic parameters (17).

The dispersion constant τ_{disp} (s) was measured as the response to step-down functions for different pumping speed and distance between the withdrawing point and the detector. Experiments were performed at room temperature. A zero dead volume junction, considered as the withdrawing point, was made by inserting 2 needles at 1 end of a 50-cm-long PE50 or PE10 catheter. Each needle was connected to a small piece of catheter and to a 3-mL syringe mounted on a computer-controlled syringe pump. The long piece of catheter was then inserted in the detector and ended in a waste container. Both syringes were filled with a nonradioactive solution, and ^{18}F -FDG was added in one of the syringes. Software was developed to operate both pumps for creating radioactivity edges. Measured edges were normalized to 1 and fitted to a simple monoexponential (18) using the software Prism 4 (GraphPad Software) to obtain τ_{disp} (s):

$$d(t) = \text{Span} \times e^{(-t \times \ln(2) / \tau_{\text{disp}})} + \text{Plateau}, \quad \text{Eq. 6}$$

where $\text{Span} + \text{Plateau}$ is the starting point and Plateau is the ending point of the exponential.

A first set of experiments was conducted with rat heparinized whole blood (3-mL samples in a 4-mL Vacutainer (Becton Dickinson) loaded with 7.2 mg K_2EDTA [EDTA is ethylenediaminetetraacetic acid]) and 3.7 MBq/mL ^{18}F -FDG for 3 different pumping speeds v (16, 62, and 250 μL/min) and 2 distances d (10 and 20 cm), for both PE10 and PE50. A second set of experiments was performed using a 400-g/L sucrose solution and 4.3 MBq/mL ^{18}F -FDG to obtain enough points to model the dispersion as a function of v and d . The sucrose concentration was selected to obtain a viscosity of 3.5 mPa·s (19), equivalent to blood viscosity (1.5–9 mPa·s depending on the hematocrit concentration (20)). Dispersion was measured with the μBC for 4 distances d (10, 20, 30, and 40 cm) at 6 different speeds v for PE50 (16, 31, 62, 125, 250, 500 μL/min) and 5 speeds for PE10 (16, 31, 62, 125, 250 μL/min). To ensure reproducibility, measurements were repeated 3 times in each configuration.

Using these parameters, the true animal blood time–activity curve $\bar{C}_a(t)$ can be computed by deconvolving the dispersion function from the measured blood time–activity curve $g(t)$ using Laplace Transforms (21):

$$\bar{C}_a(t) = g(t) + \tau_{\text{disp}} \times (dg/dt). \quad \text{Eq. 7}$$

The dispersion constant τ_{disp} can then be parameterized as follows:

$$\begin{aligned} \tau_{\text{disp}} &= Y_{\text{max}D} \times (1 - e^{-K_D \times d}) \\ &= (\text{Span}_V \times e^{-K_V \times v} + \text{Plateau}_V) \times (1 - e^{-K_D \times d}), \end{aligned} \quad \text{Eq. 8}$$

where K_D is an exponential growth rate constant, function of the distance d , and $Y_{\text{max}D}$ is the asymptotic value fitted from the starting point $\text{Span}_V + \text{Plateau}_V$ and decreasing exponentially with a rate constant K_V to Plateau_V as a function of speed v .

Animal Preparation

Animal experiments were performed in accordance with the recommendations of the in-house Ethics Committee for Animal

Experiments and the Canadian Council on Animal Care. The animals were anesthetized with isoflurane (2.5% + 1.5 L/min O₂) and cannulated with 50-cm-long catheters (PE50 for rats and PE10 for mice), filled with heparinized saline (0.9%, 50 U/mL). For arterial cannulations, a 3-way stopcock was added just before the withdrawing syringe to prevent blood from flowing up to the syringe (Fig. 3A). During animal preparation, the PET acquisition protocol was set up, and information about the experiment (animal data, injection and blood withdrawing protocols, etc.) was entered in the μ BC control software. Up to 6 time intervals with different user-defined withdrawing speeds can be set up, and the software automatically calculates the total amount of blood withdrawn, flagging any excess of 20% of the animal blood volume (~ 3.2 mL for rats and 0.4 mL for mice) (8). Withdrawing speeds down to 8 μ L/min have been used without blood clotting with PE10. At this speed, acquisitions up to 400 min for rats and 50 min for mice are possible.

After being instrumented, the animals were transported to the avalanche photodiode-based Sherbrooke PET scanner (22), and the heart was placed in the center of the camera field of view. The cannula was then inserted inside the detector unit placed next to the animal on the scanner bed, the withdrawing syringe was installed on the μ BC main unit (Fig. 3B), and the counter discrimination threshold was automatically adjusted by the software. Both the PET and counter data acquisitions were then started in synchronization, and the radiotracer administration was initiated 30 s later in the tail vein. Rats were injected with 0.5 mL of radioactive solution at a rate of 2 mL/min. A 0.35-mL saline flush was started at the end of the injection at a rate of 2 mL/min. Mice were injected with a 0.1 mL radioactive solution at a rate of 1 mL/min but no flush was made to minimize the injected volume. The μ BC control software took care of the different withdrawing speeds previously set in the protocol as the heart was imaged by the camera. PET images were reconstructed using the maximum-likelihood expectation maximization (MLEM) algorithm with 15 iterations (23).

PET Measurements

An initial animal study was performed to validate the blood time–activity curve measured by the μ BC against manual blood samples. A 241-g male Sprague–Dawley rat was cannulated in the femoral and carotid artery. The animal was injected with 37 MBq ¹⁸F-FDG. Blood was withdrawn from the femoral cannula at a rate

of 0.5 mL/min (3.2 cm/s; transit time in the detector = 1 s) for 3 min and at 50 μ L/min (0.3 cm/s; transit time = 9.5 s) for 27 min (total blood volume = 2.85 mL). The detection unit was placed 20 cm away from the cannula. Samples (~ 120 μ L) were manually taken from the carotid cannula every minute for the first 5 min and then every 5 min until the end of the acquisition and stored in preweighed tubes. At the end of the acquisition, sampling tubes were weighed again and counted in a high-sensitivity γ -well counter (Cobra II Auto-gamma; Canberra-Packard) to accurately determine samples' radioactivity concentration.

A second set of experiments was performed to investigate the measurement of MMRG in rats using the μ BC for arterial and venous blood counting. As arterial plasma radioactivity concentration is required for MMRG evaluation (3), no numeric values will be given in this study. Six female Sprague–Dawley rats, 257 ± 8 g, were cannulated in the femoral artery ($n = 3$) or in the jugular vein ($n = 3$) and injected with 34 ± 0.8 MBq ¹⁸F-FDG. For both groups, blood was withdrawn at 300 μ L/min (1.9 cm/s; transit time = 1.6 s) for 4 min and at 20 μ L/min (0.1 cm/s; transit time = 24 s) for 26 min (total blood volume = 1.72 mL). The detection unit was placed 29 ± 0.5 cm away from the cannula.

A third set of measurements was performed to illustrate the use of the μ BC for MBF evaluation in mice. Three female BALB/c mice, 20 ± 2 g, were cannulated in the carotid artery and injected with 59 ± 16 MBq ¹³N-NH₃. Blood was withdrawn at 30 μ L/min (0.8 cm/s; transit time = 3.6 s) for the entire 10-min acquisition (total blood volume = 300 μ L). The detection unit was placed 29 ± 0.5 cm away from the cannula. The blood curves were corrected for dispersion by Equation 7 using the dispersion constants measured with the sucrose solution.

Kinetic Modeling

For kinetic modeling in the rat, the blood curve measured by the μ BC was first corrected for radioactive decay and propagation delay in the catheter and smoothed to reduce statistical noise (gaussian filter, SD 2, span 5). No corrections for plasma and metabolites were made. PET tissue curves were also corrected for radioactive decay, but not for spillover, partial-volume effect, scatter, and attenuation. As PET data were in units of cps/pixel, a calibration was made by comparing PET blood data measured from a region of interest (ROI) on the left ventricle and the μ BC blood time–activity curve in the peak region. The calibration




FIGURE 3. (A) Schematic view of μ BC practical use in small-animal studies. PE50 catheter and flush are commonly used for rats; PE10 catheter without flush is typical of mouse studies. (B) Typical μ BC setup for use with Sherbrooke PET scanner shows detector unit placed near the animal on the bed and main pumping unit on left, together with an injection pump and a gas anesthesia system.




FIGURE 4. Data corrections. (A) Relationship between μBC detection efficiency K and detection threshold DI . (B) Variation of dispersion time constant τ_{disp} with distance d between animal and withdrawing point for PE50 catheter. Measured data (symbols) were fitted with monoexponential model (dashed lines).

factor was then applied to the tissue time–activity curve used in the kinetic model. For the mouse study, the scanner resolution ($2.1 \times 2.1 \times 3.1 \text{ mm}^3$ or $14 \mu\text{L}$) was not good enough to distinguish the left ventricle from the myocardium. A single ROI enclosing both structures was thus drawn and used as the tissue curve in the kinetic model.

MMRG values evaluated with ^{18}F -FDG were computed using least-square nonlinear regression with a 3-compartment and 5-parameter (K_1 , k_2 , k_3 , k_4 , and blood volume) kinetic model (3). MBF values evaluated with ^{13}N - NH_3 were computed following the same method with a 3-compartment and 4-parameter (K_1 , k_2 , k_3 , and blood volume) kinetic model (4).

RESULTS

Data Corrections

Sensitivity. The calibration curve obtained with ^{18}F is illustrated in Figure 4A and appears to fit very well to the linear relationship between DI_{min} and DI_{max} . Sensitivity parameters K_{opt} , A_{1min} , A_{2min} , and A_{max} computed using Equations 2–5 are summarized in Table 1 for different isotopes and

both catheter types. These sensitivities appear to be suitable for most pharmacokinetic PET studies in small animals.

Dispersion. All dispersion constants measured with PE50 using blood and sucrose solutions are reported in Table 2. For PE10, the dispersion constant measured with the sucrose solution ranges from $2.5 \pm 1.0 \text{ s}$ ($d = 10 \text{ cm}$, $v = 250 \mu\text{L}/\text{min}$) to $15.1 \pm 1.0 \text{ s}$ ($d = 40 \text{ cm}$, $v = 16 \mu\text{L}/\text{min}$), whereas it ranges from 1.4 s ($d = 10 \text{ cm}$, $v = 250 \mu\text{L}/\text{min}$) to 6.8 s ($d = 20 \text{ cm}$, $v = 16 \mu\text{L}/\text{min}$) with blood.

The dispersion constant obtained with blood is in good agreement, within statistical variations, with the sucrose measurements for $d = 10 \text{ cm}$. For 20 cm , blood dispersion appears statistically lower than the sucrose dispersion at 250 and $62 \mu\text{L}/\text{min}$ and higher at $16 \mu\text{L}/\text{min}$. As only one measurement is available with blood, these differences can be attributed to experimental uncertainties. Nevertheless, the sucrose solution data provide a suitable model for blood dispersion measurements.

Figure 4B gives a graphical representation of the sucrose data of Table 2 for PE50 and illustrates the exponential

TABLE 1

Optimum Absolute Sensitivities and Detection Limits Measured for 4 Common PET Radiotracers with PE50 and PE10 Catheters (Optimum Threshold, No Background Noise)

Radiotracer	$E_{\beta_{max}}$ (MeV)	Emission probability	PE50				PE10			
			K_{opt} (cps/Bq)	A_{1min} (Bq/ μL)	A_{2min} (Bq/ μL)	A_{max} (kBq/ μL)	K_{opt} (cps/Bq)	A_{1min} (Bq/ μL)	A_{2min} (Bq/ μL)	A_{max} (kBq/ μL)
^{64}Cu	0.578, 0.653	55.1%	3%	55	28	276	3.5%	203	104	1,014
^{18}F	0.634	96.7%	7.1%	23	12	116	9.6%	74	38	370
^{11}C	0.96	99.7%	21%	8	4	39	23.6%	30	15	150
^{13}N	1.199	99.8%	30%	6	3	27	34%	21	11	104

Values were measured with the μBC in use at Sherbrooke small-animal PET center in 2007 and may differ from one device to another.

TABLE 2

Dispersion Time Constant τ_{disp} (s) as Function of Withdrawing Speed and Animal-to-Detector Distance with PE50 Catheter

Pumping speed	Blood		Sucrose solution			
	10 cm	20 cm	10 cm	20 cm	30 cm	40 cm
500 $\mu\text{L}/\text{min}$			3.5 ± 0.2	5.5 ± 0.1	6.7 ± 0.3	8.0 ± 0.1
250 $\mu\text{L}/\text{min}$	5.8	6.4	5.9 ± 0.2	9.5 ± 0.2	11.0 ± 0.5	12.2 ± 0.2
125 $\mu\text{L}/\text{min}$			9.4 ± 0.4	16.4 ± 0.1	19.4 ± 0.6	22.3 ± 0.1
62 $\mu\text{L}/\text{min}$	10.7	13.9	11.1 ± 0.4	16.6 ± 0.1	20.6 ± 0.1	23.8 ± 0.3
31 $\mu\text{L}/\text{min}$			17.5 ± 0.2	23.9 ± 0.4	30.5 ± 0.2	30.3 ± 0.2
16 $\mu\text{L}/\text{min}$	23.9	43.3	24.3 ± 0.1	33.0 ± 0.2	43.6 ± 0.1	43.5 ± 0.1

Dispersion values reported for sucrose solution are mean \pm SD of 3 measurements.

relationships of Equation 8. K_D was computed for each withdrawing speed and appears quite steady for each catheter type but differs for $v \geq 62 \mu\text{L}/\text{min}$ and for $v < 62 \mu\text{L}/\text{min}$. Table 3 summarizes the parameters to be used in Equation 8 for both catheters, including K_D for each speed range. Using Equation 8 and Table 3, one can calculate the dispersion constant to be used in Equation 7 for any experimental protocol with the μBC .

PET Measurements

Blood time–activity curves obtained from manual samples and from μBC for the same animal are presented in Figure 5A. Both curves show a very good correlation, but the μBC one was much easier and faster to obtain. The μBC blood time–activity curves for rats in each group of the MMRG study are shown in Figure 5B. In the peak region, arterial sampling shows a much higher activity than venous sampling, indicating that a large fraction of the ^{18}F -FDG is absorbed in the first pass through systemic capillaries. After 200 s, the venous curve rises slightly above the arterial one, indicating some diffusion of the tracer from the tissue back to capillaries. Such patterns are typical of rapidly extracted tracers (3,24) and provide a nice demonstration of the ability of the μBC to measure accurate blood time–activity curves. The mean MMRG value computed for rats cannulated in the jugular vein was found to be underestimated by $>60\%$ in comparison with the mean MMRG value computed for rats cannulated in the femoral artery. This simple experiment stresses the importance of the blood withdrawal site for accurate measurement of the tissue metabolic rate of glucose.

TABLE 3

Parameters To Be Used in Equation 8 To Determine Dispersion Constant τ_{disp} as Function of Withdrawing Speed v and Animal-to-Detector Distance d

Parameter	PE10	PE50
Span_V	70.0	43.5
$K_V \times 10^{-2}$	4.2	1.5
Plateau_V	4.0	12.8
$K_D \times 10^{-2}$ ($v \geq 62 \mu\text{L}/\text{min}$)	6.5	3.7
$K_D \times 10^{-2}$ ($v < 62 \mu\text{L}/\text{min}$)	1.4	5.9

Figure 6A shows ^{13}N - NH_3 blood time–activity curves, without and with dispersion correction, obtained in a mouse using the μBC and compared with the time–activity curve derived from a ROI drawn on the PET images.

MBF computed without dispersion correction gave $1.12 \pm 0.17 \text{ mL}/\text{g}/\text{min}$. With dispersion correction, the MBF increased to $4.84 \pm 0.8 \text{ mL}/\text{g}/\text{min}$. As a comparison, a mean of $4.56 \text{ mL}/\text{g}/\text{min}$ was evaluated in rats with a similar experimental protocol in (25), and values from 2.5 to $4.9 \text{ mL}/\text{g}/\text{min}$ were measured in (26). This study highlights the importance of the dispersion correction for MBF measurements in mice with an external online blood counting system.

DISCUSSION

Invasiveness of online counting systems, such as the μBC , is often mentioned as a drawback, and some noninvasive blood time–activity curve measurement techniques are being developed to overcome this difficulty (5). However, these methods can hardly be applied with small animals. Similarly, accurate blood time–activity curve measurement from PET images is not always possible because of partial-volume or spillover contamination from adjacent organs in the blood pool. Furthermore, the initial investigation of new radiotracers requires direct and accurate determination of plasma and metabolite concentration of the tracer as a function of time (27). The proposed flow-through whole-blood counter allows an automated blood sampler to be inserted on the same withdrawing line (11).

Like PET scanners, most external online counting systems are based on annihilation photon detection (11,28). Although this detection principle is very reliable, it requires complex electronics and bulky detectors that necessitate cumbersome shielding. The direct detection of β -particles in photodiodes greatly simplifies the electronics and reduces the footprint next to the animal. Furthermore, as both positrons and electrons can be detected by photodiodes, the overall detection efficiency of some positron emitters—such as ^{64}Cu , for instance—can be significantly improved by measuring β^- as well as β^+ -decays. Though not tested




FIGURE 5. Blood time-activity curves corrected for radioactive decay and propagation delay obtained after bolus injection of 15 s started at 30 s, followed by 10-s flush: (A) Obtained manually and with μ BC. (B) Obtained with μ BC for arterial (light) and venous (dark) sampling (curves were smoothed with a gaussian filter (variance 2, span 5) to decrease statistical error).

in this study, the ability to detect β^- -decaying radiotracers stands out as an advantage of the proposed device as compared with annihilation detection devices. Lastly, the μ BC automated calibration process overcomes one of the significant drawbacks of most direct β^- -detection devices: the need for a different calibration factor for every isotope. This feature

alone makes the μ BC much easier to use than photon detection devices that require frequent calibrations.

Data Corrections

Sensitivity. The linear relationship between sensitivity and detection threshold in its setting range greatly simplifies




FIGURE 6. Decay corrected time-activity (TAC) curves obtained after injection of 41 MBq ^{13}N - NH_3 in 18.5-g BALB/c mouse. Bolus injection of 6 s was started at 30 s. (A) Blood time-activity curves from μ BC and from PET ROI. μ BC curve 1 is corrected for propagation delay and smoothed to decrease statistical error (gaussian filter, variance 2, span 5). μ BC curve 2 illustrates dispersion correction applied on curve 1. PET curve 3 was calculated from ROI enclosing myocardium and left ventricle. (B) Measured blood and tissue time-activity curves with blood and tissue tracer uptake extracted by kinetic modeling.

the calibration process of the μ BC. In addition to ^{18}F and PE50, this feature was also confirmed for the 3 other PET isotopes tested in this study (^{13}N , ^{11}C , ^{64}Cu) with both PE50 and PE10 catheters. The whole calibration procedure was fully integrated to the μ BC control software and takes only a few minutes. Calibration must be done for every new radioisotope before an experiment and should be repeated periodically (e.g., every month) for known radionuclides to ensure stability. However, the μ BC was used at the Sherbrooke Molecular Imaging Centre for several months, and no significant variation of sensitivity was observed. A longer-term study is needed to determine the suitable time interval for periodic calibrations.

The activities measured in the 3 animal experiments presented in this work are well within the sensitivity range of the μ BC. The typical blood peak activities of 0.5–2 kBq/ μL ^{18}F -FDG for rats and 7–12 kBq/ μL $^{13}\text{NH}_3$ for mice are 1 order of magnitude lower than the μ BC maximum count rates of 116 kBq/ μL ^{18}F -FDG with PE50 and 104 kBq/ μL $^{13}\text{NH}_3$ with PE10, respectively. As well, minimum expected activities in animals, in the range of 200–500 Bq/ μL ^{18}F -FDG for rats and 400–500 Bq/ μL $^{13}\text{NH}_3$ for mice, are 1 order of magnitude higher than the μ BC minimum detection threshold of 12 Bq/ μL ($A_{2\min}$) or 23 Bq/ μL ($A_{1\min}$) ^{18}F -FDG for PE50 and 11 Bq/ μL ($A_{2\min}$) or 21 Bq/ μL ($A_{1\min}$) $^{13}\text{NH}_3$ for mice.

On the basis of the measured sensitivities for 4 common positron-emitting radionuclides, it is expected that the μ BC will detect any isotope emitting β^+ - or β^- -particles with $E_{\beta\max} > 500$ keV and emission probability $> 50\%$. A nonexhaustive list of radioisotopes used in nuclear medicine satisfying these requirements includes ^{11}C , ^{13}N , ^{18}F , ^{64}Cu , ^{15}O , ^{68}Ga , ^{82}Rb , ^{124}I , and ^{32}P . Therefore, the μ BC could potentially find applications in areas of biomedical research other than PET.

Dispersion. According to Votaw and Shulman (17), a dispersion time constant of 10 s leads to 33% error in the measured cerebral blood flow with H_2^{15}O , whereas a 1.3-s dispersion constant leads to only 0.3% error. Dispersion in the μ BC should then be corrected for most animal studies, except for protocols leading to dispersion constants below a few seconds. According to Table 2, such low dispersions are achieved with the PE50 catheter at $v > 250$ $\mu\text{L}/\text{min}$ and $d < 20$ cm. Similarly, experiments with the PE10 catheter should not require dispersion correction with $v > 125$ $\mu\text{L}/\text{min}$ or $d < 10$ cm.

For animal studies, dispersion constants computed with Equation 8 and parameters from Table 3 can be considered as good approximations rather than absolute values, as measurements were performed with a sucrose solution at room temperature rather than whole blood at body temperature. Furthermore, occasional clotting and formation of air bubbles increased the SD on these measurements, especially with PE10 tubing because of its smaller internal diameter.

Even though manual sampling is usually considered as a reference method avoiding dispersion, it can hardly be used

to confirm the accuracy of the dispersion constants or the adopted correction method. In practice, as illustrated in Figure 5A, it is hard to obtain a sufficient number of sampling points in the peak region to draw any firm conclusions from the comparison with the μ BC curve. Blood curves obtained from PET image ROIs are also considered as a good reference with no dispersion. In the mouse study presented in this work, the PET time–activity curve had a time resolution of 5 s in the peak region, which was considered sufficient for use as a reference (Fig. 6). Nevertheless, a few issues need to be highlighted about the comparison between the μ BC and the PET time–activity curve. First, even with the 5-s sampling time period of the PET curve, the rising edge of the peak region only comprises one sample, which could lead to an underestimation of the MBF. Second, the time shift of the μ BC dispersion-corrected peak activity as compared with the PET curve could be attributed to 3 phenomena: the dispersion occurring between the target organ and the blood sampling site in the animal, an underestimation of the dispersion constant, and uncertainty in the mathematic modeling. The consequence is an underestimation of the blood peak activity in the kinetic model (Fig. 6B) and an underestimation of the MBF.

Some underestimation of the dispersion constant used to correct the mice blood curve may result from the use of a sucrose solution at room temperature rather than whole blood at body temperature. As well, differences of solubility between FDG, used for the evaluation of τ_{disp} , and NH_3 , used for this animal study, could affect dispersion. Further refinement of the experimental protocol or mathematic modeling for evaluation of the dispersion constants could provide more accurate results. Votaw and Shulman (17) proposed a double-exponential model that seems to give better results than the monoexponential model used in this study for the evaluation of the cerebral oxygen metabolic rate with $^{15}\text{O}-\text{O}_2$.

Even though further enhancements would be useful for some more demanding protocols, the dispersion was fully modeled and an automatic correction has been implemented in the control software, which together with all other corrections greatly simplifies pharmacokinetic studies.

PET Measurements

As manual sampling and μ BC data were calibrated separately, the very good agreement between the 2 curves confirms the accuracy of the μ BC automatic calibration process, even though only one point is available in the peak region of the manual sampling curve. Because better time resolution would be very hard to achieve with the manual sampling technique in small animals, no further attempt was made to achieve a better comparison of the 2 techniques in the peak region.

The discrepancy observed between MMRG measurements based on venous and arterial blood sampling has important bearings on the design of molecular imaging protocols. Arterial is more invasive than venous cannulation, making it more difficult to repeat follow-up studies on

the same animal, unless advanced cannulation techniques such as a permanent carotid catheter is used (29).

Our data show that the μ BC sensitivity is good enough to reveal subtle physiologic processes such as the difference in arterial versus venous radiotracer concentration or the second pass through the body after bolus administration of the radiotracer (Fig. 5B).

The MBF study in a mouse highlighted the fact that it is not always easy to extract the input function from an ROI placed on the blood pool in the PET image. Without the external measurement provided by the μ BC, it would not have been possible to evaluate MBF from the single ROI drawn over the whole heart, encompassing both the myocardium and the ventricular blood pool. A scanner with much better spatial resolution, sensitivity, and cardiac gating (30) is otherwise required to be able to draw 2 distinct ROIs with manageable spillover and partial-volume effects. Because of the lack of published data for mouse MBF in the literature, the values obtained in the present study were compared with rat MBF. Comparable results were obtained in spite of the fact that the generally higher metabolism in mice would have suggested a higher MBF as compared with rats.

CONCLUSION

A handy microvolumetric blood counter enabling reliable and reproducible measurement of blood time–activity curves in small animals was developed. The shielded β -detector unit is very compact and can be used in close proximity to small animals with negligible contamination from background radiation. The counter sensitivity and detection limit range from 3.5% and 104 Bq/ μ L with ^{64}Cu and PE10 catheter up to 30% and as low as 3 Bq/ μ L with ^{13}N and PE50 catheter. Such sensitivities were found to be adequate for most pharmacokinetic studies in small animals. Dispersion was fully modeled and an automatic correction was implemented in the μ BC software. Acceptable estimates of the input function were obtained in comparison with dispersion-free references, such as manual sampling or PET ROI. Selected animal studies demonstrated the accuracy of μ BC measurements when dispersion is not an issue and highlighted the importance of dispersion correction for MBF evaluation using ^{13}N -ammonia. Other studies showed the relevance of arterial sampling for MMRG measurements as well as the importance of the μ BC external measurement for MBF evaluation in mice.

The μ BC is fully automated and remote controlled by software through a flexible user-friendly interface. As a stand-alone device, the counter can be operated in conjunction with injection and flush pumps. In PET studies, the counter can be fully integrated to the Sherbrooke or LabPET scanners for managing radiotracer administration and recording blood data together with dynamic PET images. The μ BC is intended to be used with rodents but could be used as well with larger animals, such as dogs or pigs, and would

even be suitable for humans (although this has not been investigated). Because of its unique features, the μ BC should contribute to reduce staff exposure and increase throughput of pharmacokinetic studies in biomedical and pharmaceutical research.

ACKNOWLEDGMENTS

The authors thank Réjean Fontaine and Daniel Rouleau for advice on electronic design, Etienne Croteau for help on kinetic modeling, and Catherine Pepin for proof reading and suggestions. This work was part of the first author's Master's thesis. This work was supported by Canadian Institutes of Health Research (CIHR) grant MOP-15348.

REFERENCES

- Cherry SR, Gambhir SS. Use of positron emission tomography in animal research. *ILAR J*. 2001;42:219–232.
- Chatziioannou AF. Molecular imaging of small animals with dedicated PET tomographs. *Eur J Nucl Med Mol Imaging*. 2002;29:98–114.
- Phelps ME, Huang SC, Hoffman EJ, Selin C, Sokoloff L, Kuhl D. Tomographic measurement of local cerebral glucose metabolic rate in humans with (F-18)2-fluoro-2-deoxy-D-glucose: validation of method. *Ann Neurol*. 1979;6:371–388.
- Muzik O, Beanlands RS, Hutchins GD, Mangner TJ, Nguyen N, Schwaiger M. Validation of nitrogen-13-ammonia tracer kinetic model for quantification of myocardial blood flow using PET. *J Nucl Med*. 1993;34:83–91.
- Laforest R, Sharp TL, Engelbach JA, et al. Measurement of input functions in rodents: challenges and solutions. *Nucl Med Biol*. 2005;32:679–685.
- Croteau E, Benard F, Cadorette J, et al. Quantitative gated PET for the assessment of left ventricular function in small animals. *J Nucl Med*. 2003;44:1655–1661.
- Bentourkia M. Kinetic modeling of PET data without blood sampling. *IEEE Trans Nucl Sci*. 2005;52:697–702.
- Diehl KH, Hull R, Morton D, et al. A good practice guide to the administration of substances and removal of blood, including routes and volumes. *J Appl Toxicol*. 2001;21:15–23.
- Kriplani A, Stoll SP, Southekal S, et al. Noninvasive high-resolution detection of the arterial and venous input function through a PET wrist scanner. *IEEE Nuclear Science Symposium Conf Record*. 2005;4:2240–2244.
- Wu HM, Sui G, Lee CC, et al. An integrated microfluidic blood sampler for determination of blood input function in quantitative mouse microPET studies. *IEEE Nuclear Science Symposium Conf Record*. 2005;3:1658–1661.
- Lapointe D, Cadorette J, Rodrigue S, Rouleau D, Lecomte R. A microvolumetric blood counter/sampler for metabolic PET studies in small animals. *IEEE Trans Nucl Sci*. 1998;45:2195–2199.
- Yamamoto S, Tarutani K, Suga M, Minato K, Watabe H, Iida H. Development of a phoswich detector for a continuous blood-sampling system. *IEEE Trans Nucl Sci*. 2001;48:1408–1411.
- Pain F, Lanière P, Mastriopolo R, Gervais P, Hantraye P, Besret L. Arterial input function measurement without blood sampling using a β -microprobe in rats. *J Nucl Med*. 2004;45:1577–1582.
- Convert L, Morin-Brassard G, Cadorette J, et al. A microvolumetric β blood counter for pharmacokinetic PET studies in small animals. *IEEE Trans Nucl Sci*. 2007;54:173–180.
- Convert L. *Compteur sanguin μ volumétrique pour l'imagerie moléculaire chez le petit animal* [master's thesis]. Sherbrooke, Québec, Canada: Faculty of Medicine, Université de Sherbrooke; 2006.
- Knoll GF. *Radiation Detection and Measurement*. 3rd ed. New York, NY: John Wiley and Sons; 2000.
- Votaw JR, Shulman SD. Performance evaluation of the Pico-Count flow-through detector for use in cerebral blood flow PET studies. *J Nucl Med*. 1998;39:509–515.
- Iida H, Kanno I, Miura S, Murakami M, Takahashi K, Uemura K. Error analysis of a quantitative cerebral blood flow measurement using H_2^{15}O autoradiography and positron emission tomography, with respect to the dispersion of the input function. *J Cereb Blood Flow Metab*. 1986;6:536–545.

19. Mathlouthi M, Genotelle J. Rheological properties of sucrose solutions and suspensions. In: *Sucrose Properties and Applications*. London, U.K.: Blackie Academic & Professional; 1995:126–154.
20. Pirofsky B. The determination of blood viscosity in man by a method based on Poiseuille's law. *J Clin Invest*. 1953;32:292–298.
21. Vafaei M, Murase K, Gjedde A, Meyer E. Dispersion correction for automatic sampling of O-15-labeled H₂O and red blood cells. In: Myers R, Cunningham V, Bailey D, Jones T, eds. *Quantification of Brain Function Using PET*. New York, NY: Academic Press; 1996:72–75.
22. Lecomte R, Cadorette J, Richard P, Rodrigue S, Rouleau D. Design and engineering aspects of a high resolution positron tomograph for small animal imaging. *IEEE Trans Nucl Sci*. 1994;41:1446–1452.
23. Selivanov VV, Picard Y, Cadorette J, Rodrigue S, Lecomte R. Detector response models for statistical iterative image reconstruction in high resolution PET. *IEEE Trans Nucl Sci*. 2000;47:1168–1175.
24. Chiou WL. The phenomenon and rationale of marked dependence of drug concentration on blood sampling site: implications in pharmacokinetics, pharmacodynamics, toxicology and therapeutics. Part I. *Clin Pharmacokinet*. 1989;17:175–199.
25. Bentourkia M, Croteau E, Langlois R, et al. Cardiac studies in rats with ¹¹C-acetate and PET: a comparison with ¹³N-ammonia. *IEEE Trans Nucl Sci*. 2002;49:2322–2327.
26. Lecomte R, Lapointe D, Bentourkia M, Morin C, Plourde K. Cardiac PET imaging in the rat: a new tool for pharmacological studies of cardiovascular disease. *Functional Imaging in Drug Discovery and Development Conference*. Drug Information Association, Paris, France; September 30–October 2, 1999.
27. Croteau E, Benard F, Dubuc C, et al. Correction of metabolite and spillover for myocardial blood flow with ammonia in small animal PET [abstract]. *Mol Imaging Biol*. 2004;6:1P-94.
28. Kudomi N, Choi E, Yamamoto S, et al. Development of a GSO detector assembly for a continuous blood sampling system. *IEEE Trans Nucl Sci*. 2003;50:70–73.
29. Waynforth HB, Flecknell PA. *Experimental Surgical Technique in the Rat*. 2nd ed. New York, NY: Academic Press; 1992.
30. Yang Y, Rendig S, Siegel S, Newport DF, Cherry SR. Cardiac PET imaging in mice with simultaneous cardiac and respiratory gating. *Phys Med Biol*. 2005;50:2979–2989.



The Journal of
NUCLEAR MEDICINE

A New Tool for Molecular Imaging: The Microvolumetric β Blood Counter

Laurence Convert, Guillaume Morin-Brassard, Jules Cadorette, Mélanie Archambault, M'hamed Bentourkia and Roger Lecomte

J Nucl Med. 2007;48:1197-1206.

Published online: June 15, 2007.

Doi: 10.2967/jnumed.107.042606


This article and updated information are available at:
<http://jnm.snmjournals.org/content/48/7/1197>

Information about reproducing figures, tables, or other portions of this article can be found online at:
<http://jnm.snmjournals.org/site/misc/permission.xhtml>

Information about subscriptions to JNM can be found at:
<http://jnm.snmjournals.org/site/subscriptions/online.xhtml>

The Journal of Nuclear Medicine is published monthly.
SNMMI | Society of Nuclear Medicine and Molecular Imaging
1850 Samuel Morse Drive, Reston, VA 20190.
(Print ISSN: 0161-5505, Online ISSN: 2159-662X)

© Copyright 2007 SNMMI; all rights reserved.

 SOCIETY OF
NUCLEAR MEDICINE
AND MOLECULAR IMAGING