


STUDY PROTOCOL

Open Access



Precise management system for chronic intractable pain patients implanted with spinal cord stimulation based on a remote programming platform: study protocol for a randomized controlled trial (PreMaSy study)

Yuanchen Cheng^{1†}, Duo Xie^{2†}, Yan Han³, Siying Guo¹, Zhenxing Sun⁴, Linkai Jing⁴, Weitao Man⁴, Dongkang Liu⁴, Kaiyuan Yang⁴, Dan Lei⁴, Zhe Meng⁴, Huifang Zhang⁴, Guoqin Wang⁴, Weiwei Wu^{4*}, Guihuai Wang^{4*} and Yang Lu^{4,5*} 

Abstract

Background Spinal cord stimulation (SCS) is a surgical technique used in patients with chronic intractable pain, and its effectiveness and safety have been validated by multiple studies. However, to maintain an optimal and steady long-term effect is still challenging. Here, we report a new management paradigm integrating smartphone application and remote programming. Chronic pain patients with SCS implants can monitor their pain status on the phone and change stimulation parameters accordingly. The PreMaSy study is a randomized controlled trial to evaluate the clinical effectiveness and safety of this precise management system.

Methods Patients with chronic intractable pain will be screened for eligibility, and 82 participants are anticipated to be enrolled in this trial. After the electrode implantation, the stimulation effectiveness will be tested. Participants with a reduction of more than 50% in the visual analog scale (VAS) will receive implantation of an implantable pulse generator and randomized (1:1) into the experimental group or control group. All participants will be asked to take online follow-ups and complete assessments using a smartphone application. Daily pain characteristic assessments and monthly quality of life questionnaires are integrated into the App, and participants will be required to complete these assessments. The daily VAS for pain intensity will be monitored and a threshold will be set based on baseline VAS score. The interventional appointment will be scheduled once the threshold is reached. The primary outcome

[†]Yuanchen Cheng and Duo Xie have contributed equally to this work.

*Correspondence:

Weiwei Wu

weiwei.wu@btch.edu.cn

Guihuai Wang

wgha01000@btch.edu.cn

Yang Lu

yanglu@btch.edu.cn

Full list of author information is available at the end of the article



© The Author(s) 2023. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

is the health condition and quality of life assessed by the five-level EuroQol five-dimensional questionnaire (EQ-5D-5L). Utility values of EQ-5D-5L will be assessed at baseline and 1, 3, and 6 months post-operative.

Discussion The PreMaSy study aims to evaluate the effectiveness and safety of a novel App-based, patient-centered, self-assessment management system for chronic intractable pain. A randomized controlled trial is designed to test the non-inferiority of this precise management system compared to the monthly online follow-ups. It is also expected to yield valuable experiences regarding precision medicine.

Trial registration ClinicalTrials.gov NCT05761392. Registered on March 07, 2023.

Keywords Chronic intractable pain, Spinal cord stimulation, Remote programming, Mobile health, Randomized controlled trial, Precision medicine

Introduction

As a growing public health concern, chronic pain has an impact on many aspects of patients' lives. Not only does it affect physical functions and limit daily activities, but it also has a detrimental impact on work, social, and family-related consequences and decreases one's quality of life [12]. Furthermore, it creates great medical and economic burdens. According to previous reports, the prevalence of chronic pain was approximately 20% in US adults [7] and over 30% in China [53]. The resulting economic burden is therefore enormous. It is estimated that the total cost of pain in the USA in 2008 was between 560 and 635 million dollars [15].

Spinal cord stimulation (SCS) is a technique that uses implanted electrodes to directly stimulate the dorsal columns of the spinal cord. It has been approved and widely used in a variety of pain control applications, for example, failed back surgery syndrome (FBSS) [42], complex regional pain syndrome [50], radiculopathy [51], and diabetic neuropathy [36, 37]. Favorable results were demonstrated by multiple clinical trials [20, 29], confirming the effectiveness of the therapy. However, there are certain hurdles that substantially hampered the real-world applications. First, it is recommended that stimulation paradigms be individualized and modified based on real-time patient feedback in order to maximize the efficacy [38]. Considering the electrode contacts and stimulation parameters such as current, voltage, and frequency, the number of all conceivable combinations is enormous. Second, significant fluctuations are reported in long-term implantations [41]. Thus, a long-term and extensive follow-up plan is often required to maintain a stable stimulation effect. Last but not the least, the disparity of medical service assessment should be taken into account. In China, for instance, only a few cities have qualified SCS-related clinics [23], and the COVID-19 pandemic has largely restricted travel between distant places, which has aggravated this disparity.

Attempts are being made to maintain maximum and stable efficacy with minimal financial and time expenditures by utilizing telemedicine and remote programming.

Our group has reported for the first time the clinical application of remote programming for pain patients with a SCS implant using a video-based real-time remote programming system [32]. Our remote sessions allowed us to deliver remote programming operations accurately and safely, while reducing costs and providing convenience for patients.

But we wanted to take one step further. Pain characteristics differ not only across individuals but also within a single individual. For example, the pain intensity may change significantly throughout the day due to weather, mood, or sleep quality, or it may simply vary over time. But these shorter-term shifts are often ignored in both traditional and online programming sessions. Patients have to wait until scheduled follow-up sessions for major adjustments in stimulation parameters. A more precise application-based, patient-centered, self-assessment management scheme has been exploring in a broad range of diseases, for example, Parkinson's disease [1], atrial fibrillation [17], and asthma [5].

Here, we introduce for the first time a precise management system for pain patients with a SCS implant, which combines self-assessment based on a customized smartphone application (App) and remote programming based on a remote and wireless programming system. In the proposed study, patients who use the precise management system will be compared to those who take conventional online follow-ups. We intend to investigate the effectiveness of this novel system and its application in a broader population.

Methods

Trial design

The PreMaSy study is a prospective, double-arm, enrichment-enrollment, randomized-controlled trial (Clinical trial registration ID: NCT05761392) designated to evaluate the effectiveness of the precise management system for chronic intractable pain. The precise management system is a novel App-based, patient-centered, self-assessment modality designed for patients who have chronic intractable pain and are treated with remote

and wireless SCS. Compared with a classic RCT, an enrichment-enrollment trial increases the proportion of likely responders in order to maximize the differences between treatment and placebo effects, for example, in many analgetic drug trials [14, 39]. To enroll more “true” responders, the enrichment-enrollment trial shifts the randomization point, and in our instance, only after participants achieve at least a 50% reduction in pain intensity and tolerate SCS well during the test period.

The timeline of this study is divided into three distinct periods: (1) screen period, (2) test period, and (3)

follow-up period (Fig. 1). Patients experiencing chronic intractable pain will be invited to participate in our study. After signing the informed consent, participants will be screened strictly for eligibility based on the inclusion and exclusion criteria (Tables 1 and 2). During the 2-week screen period, admission arrangements will be made and thorough pre-operative examinations will be administered. In addition to collecting basic information including demographics and present and historical medical history, baseline assessments including pain characteristics and quality of life will be documented as well (Fig. 2).




Fig. 1 Flow chart of the PreMaSy study

Table 1 Inclusion criteria of the PreMaSy study

Inclusion criteria
1. Have been clinically diagnosed with chronic intractable pain for at least 3 months
2. Be unresponsive to or unsatisfied with conservative treatments, including but not limited to oral medication, nerve block, epidural corticosteroids, physical- and psychological rehabilitation therapy, and chiropractic care
3. At enrollment, aged 18 years or older
4. At enrollment, average pain intensity of at least 50 out of 100 mm on a visual analog scale
5. Be voluntary to take the trial and sign the informed consent
6. Have good compliance and ability to complete post-operative follow-ups
7. Have the basic ability to read and use a mobile phone or have a caregiver who can

Table 2 Exclusion criteria of the PreMaSy study**Exclusion criteria**

1. Have bleeding complications or coagulation disorders
2. Have mental or cognitive disorders leading to inability to complete implantation surgery or post-operative follow-up
3. Have issues with the spinal cord or vertebrates that are not suitable for implantation surgery
4. Have systemic active infections or local infections at the anticipated surgery area
5. Be pregnant, breast-feeding, plan to be pregnant, or unwilling to use contraceptive methods
6. Have metastatic tumors or untreated malignancies
7. Have a life expectancy of less than 1 year
8. Have already provided with a medication pump and/or other implanted devices
9. Require the use of MRI and/or thermo-therapy
10. Be heavily addicted to alcohol and/or drugs
11. Have an improvement in VAS of less than 50% or unable to tolerance SCS during the test period
12. Unable to complete long-term online follow-ups because of hardware issues such as internet, mobile phones, and so on

	STUDY PERIOD										
	Screen Period	Test Period			Follow-up Period						
	Baseline	Electrode implantation	Allocation	IPG implantation	1-month follow-up	2-month follow-up	3-month follow-up	4-month follow-up	5-month follow-up	6-month follow-up	Daily assessment
TIMEPOINT	Day -28 ~ -14	Day -10 (± 4 days)	Day -1	Day 0	Day 30 (± 5 days)	Day 60 (± 5 days)	Day 90 (± 5 days)	Day 120 (± 5 days)	Day 150 (± 5 days)	Day 180 (± 5 days)	Day 1 ~ 180
ENROLLMENT:											
Eligibility screen	X										
Informed consent	X										
Demographics	X										
Present medical history	X										
Historical medical history	X										
Drug history	X										
Telemedicine assessment	X										
Health economics	X				X	X	X	X	X	X	
Vital signs	X										
Physical examination	X										
Laboratory tests	X										
Spine MRI	X										
Chest X-ray	X										
Surgery area X-ray		X		X							
Allocation			X								
INTERVENTIONS:											
PreMaSy group					←	→	→	→	→	→	X
Control group					X		X			X	
ASSESSMENTS:											
Programming		X		X	X	X	X	X	X	X	X
Pain intensity	X	X		X	X	X	X	X	X	X	X
Time-pain intensity curve	X	X		X							X
Pain map	X	X		X							X
EQ-5D-5L	X				X	X	X	X	X	X	
Pittsburgh sleep quality index	X				X	X	X	X	X	X	
Patient's global impression of change					X	X	X	X	X	X	
Clinical global impression-improvement					X	X	X	X	X	X	
Impedance tests		X		X	X	X	X	X	X	X	X
Parameter adjustments		X		X	X	X	X	X	X	X	
Adverse events record		X		X	X	X	X	X	X	X	X
Participants' degree of satisfaction to app					X	X	X	X	X	X	
Investigators' degree of satisfaction to app					X	X	X	X	X	X	

Fig. 2 Standard Protocol Items: Recommendations for Interventional Trials (SPIRIT) figure

Participants will then undertake the SCS electrode implantation and enter the 10-day (± 4) test period. Pain intensity will be monitored during this time. Only participants who have an improvement of no less than 50% in visual analog scale (VAS) will undertake the implantable pulse generator (IPG) implantation. After the IPG implantation, participants will be randomized (1:1) into one of the two groups: the experimental group (EG) with PreMaSy or the control group (CG) with conventional follow-ups. The randomization will be generated by computers and concealed from the recruitment team and data analysis team. The participants and their physicians, however, will not be blinded from their own randomization results, because the consequent assessments are dependent on which group they are in. The detailed assessments of the two groups are shown in Fig. 2. Participants in CG will take follow-ups only at 1, 3, and 6 months post-operative, while participants in EG will take additional monthly quality of life assessments and daily pain characteristics assessments, which can be done in a customized App. These results will be automatically monitored and timely interventional appointment will be scheduled whenever the intervention threshold is reached.

The PreMaSy study protocol was written in accordance with the Standard Protocol Items: Recommendations for Interventional Trials Statement (SPIRIT). A SPIRIT checklist is included in Additional file 1. The trial will be carried out according to the principles of the Declaration of Helsinki (Edinburgh version, 2000).

Participants

Study population

The study will be primarily conducted in the Neurosurgery Center of Beijing Tsinghua Changgung Hospital. Additional centers may be added as the study progresses. The protocol has been approved by the Clinical Trial Ethics Committee of Beijing Tsinghua Changgung Hospital (approval number: 22442-4-02). It is anticipated that the enrollment will end prior to September 2024. And invitations will be sent to our clinic patients experiencing chronic intractable pain during that time.

Inclusion criteria

The inclusion criteria are shown in Table 1.

Exclusion criteria

The exclusion criteria are shown in Table 2.

App design

Our research team in cooperation with Beijing Pinchi Software Development Co., Ltd. developed an App that integrates the wireless control of the stimulation device

and the follow-up assessments. The App can connect wirelessly to the IPG device based on our previously developed remote and wireless SCS system [32]. Four sets of personalized stimulation parameters will be programmed to optimize the analgesic effect and stored in the App before participants are discharged. Participants will be given complete control over the stimulation, including turning it on and off and changing the parameter set in use according to their preferences.

The assessment module was designed as a crucial element in the precise management system. Four major sections are included: (1) daily assessments, (2) monthly assessments, (3) hardware-related issues, and (4) satisfaction with App usage.

The daily assessment section primarily focuses on the evaluation of multiple pain characteristics, including intensity, temporal, and spatial features. Pain intensity is measured by a VAS slider customized to mobile phone screen size (Fig. 3A) [44]. To measure the temporal features, we designed a hand-drawn time-intensity curve (Fig. 3B). The *x*-axis of the curve represents 24 h in a day, while the *y*-axis represents the VAS score of pain intensity. The participants are asked to recall their pain experience from the previous day and draw a curve to depict the fluctuations in pain intensity over time. As for the spatial features, we referred to the pain area drawing in the Brief Pain Inventory (BPI) and designed a version that is touch-screen compatible. In order to standardize the pain map, the front and rear body projections were separated according to the dermatome pattern. However, the pattern was too detailed to be identified accurately on a smartphone display. Therefore, we consolidated certain consecutive dermatomes and generated 25 gross selection areas (Fig. 3C, D).

The monthly assessment section, on the other hand, focuses on the long-term pain-related quality of life. The Five-Level EuroQol Five-Dimensional Questionnaire (EQ-5D-5L) is a widely used instrument for assessing general health that consists of mobility, self-care, usual activities, pain/discomfort, and anxiety/depression [21]. We used the published value set for China to evaluate various health statuses [34]. The EQ-VAS is used as well for the global health condition. Sleep quality is assessed by the Pittsburgh Sleep Quality Index (PSQI) with 18 items covering seven categories [4]. Patients' Global Impression of Change (PGIC) and Clinical Global Impressions-Improvements (CGI-I) are two seven-level questionnaires that measure the perceptions of the improvements in the participants' overall health conditions, respectively from their own [45] and their physicians [3].

Impedance tests, parameter adjustments, adverse events (AEs), and other hardware-related issues are




Fig. 3 The daily assessments designed for the App. **A** Slider of VAS for general pain intensity. The current position indicates 5.0 out of 10.0. **B** Time-intensity curve for pain temporal characteristics. The x-axis depicts 24 h in a day in 3-h increments, and the y-axis depicts the pain intensity. The red curve simulates a patient's hand-drawn curve. **C, D** Simplified pain map based on the dermatome pattern. There are a total of 25 pairs of body area presented, including 12 front (**C**) and 13 back (**D**). The indicators highlighted in blue simulate selections from an imaginative candidate

collected. Participants and their physicians are also asked to take satisfaction surveys based on their App usage experiences and encouraged to provide feedback. All aforementioned assessment modules will be activated automatically when the profile of a participant is established, and in order to enhance compliance, notifications of incomplete assessments will be sent at a set time.

Interventions

During the test period, electrode implantation and IPG implantation will be undertaken in patients who qualify. After surgery, stimulation parameters will be optimized for analgesia. In our prior work, we have described the implantation surgery, stimulation device, and programming procedure in detail [31]. The aforementioned App will be utilized for subsequent follow-ups and pain management. Upon enrollment, an App account will be assigned to each participant. They will be instructed to activate the account and fill up the profile. As soon as they undergo the implantation procedure and are randomly assigned to either EG or CG, the App will only display contents exclusively to their group.

For participants in CG, they will be scheduled for follow-ups at conventional intervals, i.e., 1, 3, and 6 months post-operative. They can only access monthly assessments during the corresponding months. For

participants in EG, on the other hand, the precise management system will be implemented. In addition to the standard follow-ups, they will be asked to complete every monthly assessment throughout the 6-month follow-up period. They will have access to the daily assessments in which they will be instructed to record and track their pain characteristics. Pain intensity VAS data collected after the electrode implantation and prior to the IPG implantation will be averaged and served as the baseline. The interventional threshold will be a minimum clinically significant difference (MCS D) of 20 mm (out of 100 mm VAS) [27] added to the baseline value. The system will monitor each participant's daily VAS score automatically. If the score surpasses the threshold three times in a row, an alarm will be sent to their physicians and a remote programming appointment will be scheduled.

To improve adherence, the background system will monitor the assessments' completion status, and specialists will be allocated to inquire about the incompleteness. Other concomitant care including medications are only permitted by approval of investigators. Both CG and EG participants have the same access to report hardware-related issues and satisfaction with App usage. In the case of adverse events (AEs) or unexpected worsening, they may also schedule an urgent appointment.

Outcomes

Primary outcome

As previously stated, EQ-5D-5L is a questionnaire which measures five dimensions. It depicts different health conditions as five-digit numbers corresponding to unique utility values. The primary outcome of the PreMaSy study is the change of EQ-5D-5L utility values between CG and EG at the 6-month follow-up compared to baseline values. After passing the eligibility screening, they will be required to fill up an EQ-5Q-5L questionnaire based on their pain experience prior to enrollment, and this value will serve as the baseline. The primary outcome formula is:

$$(EQ_{EG6} - EQ_{EGbase}) - (EQ_{CG6} - EQ_{CGbase})$$

Secondary outcomes and safety outcomes

The differences in EQ-5D-5L utility values at 1- and 3-month follow-ups compared to baseline will be included in the secondary outcomes. Other secondary outcomes include changes in pain intensity VAS, time-intensity curve, pain map, EQ-VAS score, PSQI, PGIC, CGI-I, and satisfaction with App usage. The occurrence rate of AEs and severe AEs (SAEs) will be used as safety outcomes.

Statistical methods

Sample size

The primary comparison in this study is the difference between EG and CG in quality of life score measuring by EQ-5D-5L. The EQ-5D-5L difference is assumed to be 0.53 with a standard deviation (SD) of 0.69 based on previous studies into the effectiveness of App-based therapy for chronic pain [40]. A sample size of 82 (41 per group) is calculated by the following formula:

$$n = \frac{(1+r)(Z_{\alpha} + Z_{1-\beta})^2 \sigma^2}{r(|D| + \Delta)^2}$$

Z is the quantile of the standard normal distribution, α is the first-class error level, β is the second-class error level for the statistical test, and r is the allocation ratio, which is 1:1 in this equal allocation trial. Here, to detect the significant difference at a significant level of 0.05 with 80% power using a two-tailed t test, Z_{α} and $Z_{1-\beta}$ are determined as 1.96 ($\alpha = 0.05$) and 0.84 ($\beta = 0.2$), respectively. σ is the expected SD of the outcome, and $|D|$ is the expected mean difference of the two groups. The non-inferiority limit is set at 10% [10], and Δ is 0.053. The sample size per group is calculated with the above formula, and the given value is 22. With the

assumption of a 60% “true” responding rate and a 10% possible attrition rate, the final calculation of the sample size is 41 pairs ($N = 82$).

Primary and additional analyses

The primary effectiveness analysis is the 6-month non-inferiority of the within-group difference between CG and EG for the overall health-related quality of life measured by EQ-5D-5L. Non-inferiority will be tested between the two groups at a one-sided significance of 0.05, and if the non-inferiority of the primary endpoint was achieved, superiority will be tested at a one-sided significance of 0.025.

For secondary endpoints, non-inferiority will be also tested with a 10% non-inferiority margin and a one-sided significance of 0.05. Secondary endpoints include changes from baseline of EQ-5D-5L scores in 1 and 3 months, VAS scores, time-intensity curves, pain map, EQ-VAS scores, PSQI, PGIC, CGI-I, and satisfaction with App usage. Continuous variables will be summarized using means and standard error, and categorical variables will be summarized using frequency distributions. Parametric tests (e.g., t test and ANOVA) will be performed with data of normal distributions. Non-parametric tests (e.g., Wilcoxon rank-sum test and Mann-Whitney test) will be performed if parametric tests are not indicated. All statistical analyses will be performed in a validated statistical software package (e.g., SAS or SPSS).

Data collection and management

Each participant will be given a unique account number for the App. During hospitalization, all baseline data will be collected, and after the implantation, all participants will be instructed by qualified staff to complete the daily pain assessments. Additional instructions will be provided before the discharge. All data will be uploaded and stored in a cloud-based internet platform, and the transmission will be encrypted to ensure security and confidentiality. All the staff participating in the trial must complete the full training course, which includes participant enrollment, assessment completion on the mobile App, surgery technique, programming demonstration, database demonstration, and CRF completion. Participants and at least one caregiver should be familiar with the mobile App and remote programming. Initial data monitoring will be performed by the investigator himself/herself to ensure authenticity and completeness. Then the certified clinical research associate (CRA) will monitor, audit, and assure compliance with the protocol, GCP, and regulatory requirements during data collection.

Discussion

SCS is one of the surgical approaches for patients with chronic intractable pain. It is sometimes seen as “the last resort therapy” for those who do not respond well to medications or other non-surgical treatments. Since the first implanted SCS system in 1967 [49], significant development has been established within the past six decades.

Conventional frequency SCS (CF-SCS) delivers a frequency of 10 to 300 Hz. It targets the spinal inhibitory GABA system and usually generates paresthesias [16]. High-frequency SCS (HF-SCS) delivers a frequency of 5 to 10 kHz, and the mechanism is similar to CF-SCS but more robust. Multiple studies have verified the efficacy and safety of HF-SCS in relieving chronic back and leg pain (CBLP) [33, 48, 52] and painful diabetic neuropathy (PDN) [46]. Results from a 24-month RCT indicated the superiority of HF-SCS over CF-SCS in long-term therapy [25]. Another emerging waveform is the non-tonic burst stimulation. For burst programming, five pulses are delivered within 1 ms at a frequency of 40 Hz. Burst SCS is also safe and effective, and it has been proven to be superior to tonic CF-SCS in the treatment of chronic pain [8, 22]. There are other new current delivery methods that adjust stimulation dynamically according to the neurophysiological response to SCS, for example, evoked compound action potential (ECAP) SCS [11] and closed-loop [36, 37].

There are still challenges in the SCS field. The first one is to optimize the efficacy, which requires the personalization of stimulation paradigms according to each patient's need. But the efficacy can be affected by multiple variables, including but not limited to stimulation parameters, lead location, patient anatomy, and pain pattern [20]. Personalization could be quite difficult due to the intricacy of the stimulation paradigm and the heterogeneity of the patients. The second one is to sustain efficacy. Long-term efficacy has been demonstrated in several studies [25]. However, post-operative fluctuations due to electrode migration, electrode malposition, incision site infection, or loss of therapeutic effect were reported [29, 43]. In addition, patients often suffer fluctuations throughout the day [41], which significantly lowers their quality of life [18]. The third one is the disparity in patient access to SCS-related medical services. It is shown that economically developed regions are more likely to have better chronic pain diagnosis and treatment and vice versa [23]. This disparity makes it more difficult and costly for patients who live far from a SCS clinic to receive therapy.

To address these challenges, efforts were made in the developments of neurostimulation technologies [9]. Advances in Bluetooth technology allow wireless communication between leads and IPG, enabling remote

programming and device miniaturization. The feasibility of remote programming is validated by a randomized trial that remotely instructed nonexpert individuals to program the neuromodulation device successfully [38]. Telemedicine is also a new and promising technology that provides remote healthcare service to rural or medically underserved regions [35]. It is considered as an alternative to the conventional outpatient clinic since it improves equity of access to healthcare while also increasing the consistency and quality of service [6]. Combined with telemedicine and remote programming, geographical barriers can be broken and patients can take programming sessions in their homes [32]. In the meantime, eTools and smartphone Apps are targeting more aspects including health care, education, and self-monitoring [2]. Since the COVID-19 pandemic, access to traditional face-to-face service has been largely limited, which in turn promotes transformation toward telemedicine to meet the increasing need for remote programming [26, 30].

Efforts on developing such Apps have already been made in the management of chronic pain [13, 28]. However, the use of SCS was not covered in these studies. In our previous studies, we demonstrated the high demands for remote programming among patients with chronic intractable pain in China [19] and established the safety and effectiveness of a remote wireless programming system [32]. By integrating daily and monthly assessments into a customized smartphone App, we aim to establish an App-based, patient-centered, self-assessment precise management system.

In our customized App, information on patient education and assessments of semiological monitoring are provided as basic functions, which is consistent with other Apps [13, 28]. For semiological monitoring, daily and monthly assessments are provided. Daily assessments of pain focus on three different perspectives, namely intensity, temporal, and spatial characteristics. Monthly assessments, on the other hand, focus on the general health-related quality of life. These assessments instruct participants to focus on their pain experiences and how they alter with stimulation. Daily fluctuations will be captured, and interventions will be delivered in time based on this information. In order to maximize the potential of smartphone Apps [47], participants are also provided with a selection of stimulation paradigms, including two CF-SCS sets and two HF-SCS sets. A real-world investigation indicated that patients may have different stimulation paradigms preferences [24].

There are some limitations in this study. The primary objective is to determine whether this new precise management system is noninferior to conventional online follow-ups. However, the evidence of App effectiveness

is still absent. The effect due to the different intensities of App usage cannot be fully interpreted. Another limitation is that the intervention cannot be blinded because it makes direct changes in patients' lifestyle and relies on their feedbacks. In addition, the overall duration of follow-ups is only 6 months, which is insufficient for exploring long-term consequences and reducing placebo effects.

Trial status

Protocol version 1.3 (March 25, 2023). The recruitment began in September 2022 and is expected to be completed in September 2024.

Abbreviations

AE(s)	Adverse event(s)
App(s)	Application(s)
BPI	Brief Pain Inventory
CBLP	Chronic back and leg pain
CF-SCS	Conventional frequency spinal cord stimulation
CG	Control group
CGI-I	Clinical Global Impressions-Improvements
CRA	Clinical research associate
CRF	Case report form
EACP	Evoked compound action potential
EG	Experimental group
EQ-5D-5L	Five-Level EuroQol Five-Dimensional Questionnaire
FBSS	Failed back surgery syndrome
GCP	Good Clinical Practice
HF-SCS	High-frequency spinal cord stimulation
IPG	Implantable pulse generator
MCSD	Minimum clinically significant difference
PDN	Painful diabetic neuropathy
PGIC	Patients' Global Impression of Change
PSQI	Pittsburgh Sleep Quality Index
RCT	Randomized controlled trial
SAE(s)	Severe adverse event(s)
SCS	Spinal cord stimulation
SD	Standard deviation
VAS	Visual analog scale

Supplementary Information

The online version contains supplementary material available at <https://doi.org/10.1186/s13063-023-07595-4>.

Additional file 1. Reporting checklist for protocol of a clinical trial.

Acknowledgements

We would like to thank the entire team of researchers for their enthusiasm and insightful thoughts. We would also like to thank all the nurses and staff in the hospital.

Authors' contributions

All authors have made substantial contributions to the design and concept of this study. YC and YH drafted the work under the supervision of GW^a and YL. YC, DX, and YL designed the App used in the work. All authors have made substantively revised the manuscript and approved the submitted version. SG, ZS, LJ, WM, DL^b, KY, DL^c, ZM, HZ, GW^m, WW, GW^a, and YL were involved in the collection and analysis of the data.

Funding

This work was supported by the Beijing Nova Program (Z21100002121050, 20220484216) and Tsinghua Precision Medicine Foundation (LC201906).

Availability of data and materials

This is a study protocol only. The final trial dataset will be kept strictly confidential and can only be accessed by members of the trial team.

Declarations

Ethics approval and consent to participate

The protocol of the PreMaSy study has been approved by the Clinical Trial Ethics Committee of Beijing Tsinghua Changgung Hospital (approval number: 22442-4-02) on September 11, 2022. Any modifications to the protocol will be discussed by all trial investigators and sent for approval by the Ethics Committee. All participants will be asked for intentions to participate, and written informed consent will be obtained from all participants.

Consent for publication

This protocol does not contain any data from an individual person. The results will be reported at conferences and in peer-reviewed publications.

Competing interests

The authors declare that they have no competing interests.

Author details

¹School of Medicine, Tsinghua University, Beijing, China. ²Air Force Medical Center PLA, Beijing, China. ³Department of Neurology Xuanwu Hospital, Capital Medical University, Beijing, China. ⁴Department of Neurosurgery, Beijing Tsinghua Changgung Hospital, School of Clinical Medicine, Tsinghua University, Beijing, China. ⁵Institute for Precision Medicine, Tsinghua University, Beijing, China.

Received: 12 April 2023 Accepted: 18 August 2023

Published online: 11 September 2023

References

- Bot BM, Suver C, Neto EC, Kellen M, Klein A, Bare C, ... Trister AD. The mPower study, Parkinson disease mobile data collected using ResearchKit. *Scientific Data*. 2016;3(1):160011. <https://doi.org/10.1038/sdata.2016.11>.
- Boulos M, Wheeler S, Tavares C, Jones R. How smartphones are changing the face of mobile and participatory healthcare: an overview, with example from eCAALYX. *Biomed Eng Online*. 2011;10(1):24. <https://doi.org/10.1186/1475-925x-10-24>.
- Busner J, Targum SD. The clinical global impressions scale: applying a research tool in clinical practice. *Psychiatry (Edmont)*. 2007;4(7):28–37.
- Buyse DJ, Reynolds CF 3rd, Monk TH, Berman SR, Kupfer DJ. The Pittsburgh Sleep Quality Index: a new instrument for psychiatric practice and research. *Psychiatry Res*. 1989;28(2):193–213. [https://doi.org/10.1016/0165-1781\(89\)90047-4](https://doi.org/10.1016/0165-1781(89)90047-4).
- Chan Y-FY, Wang P, Rogers L, Tignor N, Zweig M, Hershman SG, ... Schadt EE. The Asthma Mobile Health Study, a large-scale clinical observational study using ResearchKit. *Nat Biotechnol*. 2017;35(4):354–362. <https://doi.org/10.1038/nbt.3826>.
- Craig J, Patterson V. Introduction to the practice of telemedicine. *J Telemed Telecare*. 2005;11(1):3–9. <https://doi.org/10.1177/1357633x0501100102>.
- Dahlhamer J, Lucas J, Zelaya C, Nahin R, Mackey S, Debar L, ... Helmick C. Prevalence of chronic pain and high-impact chronic pain among adults—United States, 2016. *MMWR Morb Mortal Wkly Rep*. 2018;67(36), 1001–1006. <https://doi.org/10.15585/mmwr.mm6736a2>.
- Deer T, Slavin KV, Amirdelfan K, North RB, Burton AW, Yearwood TL, ... Wilson D. Success using neuromodulation with BURST (SUNBURST) study: results from a prospective, randomized controlled trial using a novel burst waveform. *Neuromodulation*. 2018;21(1), 56–66. <https://doi.org/10.1111/ner.12698>.
- Deer TR, Krames E, Mekhail N, Pope J, Leong M, Stanton-Hicks M, ... Neuromodulation Appropriateness Consensus C. The appropriate use of neurostimulation: new and evolving neurostimulation therapies and applicable treatment for chronic pain and selected disease states. *Neuromodulation Appropriateness Consensus Committee. Neuromodulation*. 2014;17(6):599–615. discussion 615. <https://doi.org/10.1111/ner.12204>.

10. Deer TR, Levy RM, Kramer J, Poree L, Amirdelfan K, Grigsby E, ... Mekhail N. Dorsal root ganglion stimulation yielded higher treatment success rate for complex regional pain syndrome and causalgia at 3 and 12 months: a randomized comparative trial. *Pain*. 2017;158(4):669–681. <https://doi.org/10.1097/j.pain.0000000000000814>.
11. Duarte RV, Soliday N, Leitner A, Taylor RS. Health-related quality of life associated with pain health states in spinal cord stimulation for chronic neuropathic pain. *Neuromodulation*. 2021;24(1):142–9. <https://doi.org/10.1111/ner.13267>.
12. Duenas M, Ojeda B, Salazar A, Mico JA, Failde I. A review of chronic pain impact on patients, their social environment and the health care system. *J Pain Res*. 2016;9:457–67. <https://doi.org/10.2147/jpr.s105892>.
13. Fu MR, Axelrod D, Guth AA, Scagliola J, Rampertaap K, El-Shammaa N, ... Wang Y. A web- and mobile-based intervention for women treated for breast cancer to manage chronic pain and symptoms related to lymphedema: results of a randomized clinical trial. *JMIR Cancer*. 2022;8(1):e29485. <https://doi.org/10.2196/29485>.
14. Furlan AD, Chaparro LE, Irvin E, Mailis-Gagnon A. A comparison between enriched and nonenriched enrollment randomized withdrawal trials of opioids for chronic noncancer pain. *Pain Res Manage*. 2011;16(5):337–51. <https://doi.org/10.1155/2011/465281>.
15. Gaskin DJ, Richard P. The economic costs of pain in the United States. *J Pain*. 2012;13(8):715–24. <https://doi.org/10.1016/j.jpain.2012.03.009>.
16. Geurts JW, Joosten EA, van Kleef M. Current status and future perspectives of spinal cord stimulation in treatment of chronic pain. *Pain*. 2017;158(5):771–4. <https://doi.org/10.1097/j.pain.0000000000000847>.
17. Guo Y, Chen Y, Lane DA, Liu L, Wang Y, Lip GYH. Mobile health technology for atrial fibrillation management integrating decision support, education, and patient involvement: mAF app trial. *Am J Med*. 2017;130(12):1388–1396 e1386. <https://doi.org/10.1016/j.amjmed.2017.07.003>.
18. Hagedorn JM, Romero J, Ha CT, D'Souza RS. Patient satisfaction with spinal cord stimulation and dorsal root ganglion stimulation for chronic intractable pain: a systematic review and meta-analysis. *Neuromodulation*. 2022;25(7):947–55. <https://doi.org/10.1016/j.neurom.2022.04.043>.
19. Han Y, Lu Y, Wang D, Ran M, Ren Q, Xie D, ... Wang JJ. The use of remote programming for spinal cord stimulation for patients with chronic pain during the COVID-19 outbreak in China. *Neuromodulation*. 2021;24(3):441–447. <https://doi.org/10.1111/ner.13382>.
20. Hatheway JA, Mangal V, Fishman MA, Kim P, Shah B, Vogel R, ... Kelly MT. Long-term efficacy of a novel spinal cord stimulation clinical workflow using kilohertz stimulation: twelve-month results from the Vectors study. *Neuromodulation*. 2021;24(3):556–565. <https://doi.org/10.1111/ner.13324>.
21. Herdman M, Gudex C, Lloyd A, Janssen M, Kind P, Parkin D, ... Badia X. Development and preliminary testing of the new five-level version of EQ-5D (EQ-5D-5L). *Qual Life Res*. 2011;20(10):1727–1736. <https://doi.org/10.1007/s11136-011-9903-x>.
22. Hou S, Kemp K, Grabis M. A systematic evaluation of burst spinal cord stimulation for chronic back and limb pain. *Neuromodulation*. 2016;19(4):398–405. <https://doi.org/10.1111/ner.12440>.
23. Jiang Y, Xu T, Mao F, Miao Y, Liu B, Xu L, ... Fan B. The prevalence and management of chronic pain in the Chinese population: findings from the China Pain Health Index (2020). *Population Health Metrics*. 2022;20(1). <https://doi.org/10.1186/s12963-022-00297-0>.
24. Kallewaard JW, Paz-Solis JF, De Negri P, Canos-Verdecho MA, Belaid H, Thomson SJ, ... Jain R. Real-world outcomes using a spinal cord stimulation device capable of combination therapy for chronic pain: a European, multicenter experience. *J Clin Med*. 2021;10(18). <https://doi.org/10.3390/jcm10184085>.
25. Kapural L, Yu C, Doust MW, Gliner BE, Vallejo R, Sitzman BT, ... Burgher AH. Comparison of 10-kHz high-frequency and traditional low-frequency spinal cord stimulation for the treatment of chronic back and leg pain: 24-month results from a multicenter, randomized, controlled pivotal trial. *Neurosurgery*. 2016;79(5):667–677. <https://doi.org/10.1227/NEU.0000000000001418>.
26. Keesara S, Jonas A, Schulman K. COVID-19 and health care's digital revolution. *New England Journal of Medicine*. 2020;382(23):e82. <https://doi.org/10.1056/nejmp2005835>.
27. Kelly AM. The minimum clinically significant difference in visual analogue scale pain score does not differ with severity of pain. *Emerg Med J*. 2001;18(3):205–7. <https://doi.org/10.1136/emj.18.3.205>.
28. Kerckhove N, Delage N, Cambier S, Cantagrel N, Serra E, Marcaillou F, ... Authier N. eDOL mHealth app and web platform for self-monitoring and medical follow-up of patients with chronic pain: observational feasibility study. *JMIR Form Res*. 2022;6(3):e30052. <https://doi.org/10.2196/30052>.
29. Kumar K, Taylor RS, Jacques L, Eldabe S, Meglio M, Molet J, ... North RB. The effects of spinal cord stimulation in neuropathic pain are sustained: a 24-month follow-up of the prospective randomized controlled multicenter trial of the effectiveness of spinal cord stimulation. *Neurosurgery*. 2008;63(4):762–770. <https://doi.org/10.1227/01.NEU.0000325731.46702.D9>. discussion 770.
30. Lo Bianco G, Papa A, Schatman ME, Tinnirello A, Terranova G, Leoni MLG, ... Mercadante S. Practical advices for treating chronic pain in the time of COVID-19: a narrative review focusing on interventional techniques. *J Clin Med*. 2021;10(11). <https://doi.org/10.3390/jcm10112303>.
31. Lu Y, Mao P, Wang G, Tao W, Xiong D, Ma K, ... Li L. Spinal cord stimulation for chronic intractable trunk or limb pain: study protocol for a Chinese multicenter randomized withdrawal trial (CITRIP study). *Trials*. 2020;21(1):834. <https://doi.org/10.1186/s13063-020-04768-3>.
32. Lu Y, Xie D, Zhang X, Dong S, Zhang H, Yu B, ... Li L. Management of intractable pain in patients with implanted spinal cord stimulation devices during the COVID-19 pandemic using a remote and wireless programming system. *Front Neurosci*. 2020;14:594696. <https://doi.org/10.3389/fnins.2020.594696>.
33. Luecke T, Edgar D, Huse D. 10 kHz spinal cord stimulation for the treatment of chronic back and/or leg pain: summary of clinical studies. *SAGE Open Medicine*. 2020;8:205031212095136. <https://doi.org/10.1177/2050312120951369>.
34. Luo N, Liu G, Li M, Guan H, Jin X, Rand-Hendriksen K. Estimating an EQ-5D-5L value set for China. *Value Health*. 2017;20(4):662–9. <https://doi.org/10.1016/j.jval.2016.11.016>.
35. Marcin JP, Ellis J, Mawis R, Nagrampa E, Nesbitt TS, Dimand RJ. Using telemedicine to provide pediatric subspecialty care to children with special health care needs in an underserved rural community. *Pediatrics*. 2004;113(1 Pt 1):1–6. <https://doi.org/10.1542/peds.113.1.1>.
36. Mekhail N, Levy RM, Deer TR, Kapural L, Li S, Amirdelfan K, ... Poree L. Long-term safety and efficacy of closed-loop spinal cord stimulation to treat chronic back and leg pain (Evoke): a double-blind, randomised, controlled trial. *Lancet Neurol*. 2020;19(2):123–134. [https://doi.org/10.1016/s1474-4422\(19\)30414-4](https://doi.org/10.1016/s1474-4422(19)30414-4).
37. Mekhail NA, Argoff CE, Taylor RS, Nasr C, Caraway DL, Gliner BE, ... Brooks ES. High-frequency spinal cord stimulation at 10 kHz for the treatment of painful diabetic neuropathy: design of a multicenter, randomized controlled trial (SENZA-PDN). *Trials*. 2020;21(1):87. <https://doi.org/10.1186/s13063-019-4007-y>.
38. Mendez I, Song M, Chiasson P, Bustamante L. Point-of-care programming for neuromodulation: a feasibility study using remote presence. *Neurosurgery*. 2013;72(1):99–108. <https://doi.org/10.1227/NEU.0b013e318276b5b2>. (discussion 108).
39. Moore RA, Wiffen PJ, Eccleston C, Derry S, Baron R, Bell RF, ... Straube S. Systematic review of enriched enrolment, randomised withdrawal trial designs in chronic pain: a new framework for design and reporting. *Pain*. 2015;156(8):1382–1395. <https://doi.org/10.1097/j.pain.000000000000088>.
40. Morcillo-Muñoz Y, Sánchez-Guarnido AJ, Calzón-Fernández S, Baena-Parejo I. Multimodal chronic pain therapy for adults via smartphone: randomized controlled clinical trial. *J Med Int Res*. 2022;24(5):e36114. <https://doi.org/10.2196/36114>.
41. Nagel SJ, Hsieh J, Machado AG, Frizon LA, Howard MA 3rd, Gillies GT, Wilson S. Biomarker optimization of spinal cord stimulation therapies. *Neuromodulation*. 2021;24(1):3–12. <https://doi.org/10.1111/ner.13252>.
42. North RB, Kidd D, Shipley J, Taylor RS. Spinal cord stimulation versus reoperation for failed back surgery syndrome: a cost effectiveness and cost utility analysis based on a randomized, controlled trial. *Neurosurgery*. 2007;61(2):361–8. <https://doi.org/10.1227/01.Neu.0000255522.42579.Ea>. (discussion 368–369).
43. North RB, Kidd DH, Farrokhi F, Piantadosi SA. Spinal cord stimulation versus repeated lumbosacral spine surgery for chronic pain: a randomized, controlled trial. *Neurosurgery*. 2005;56(1):98–106. <https://doi.org/10.1227/01.neu.0000144839.65524.e0>. (discussion 106–107).
44. Park S, Xiong S. The effect of slider design and length on user performance and preference of smartphone versions of the visual analogue scale. *Appl Ergon*. 2021;97:103521. <https://doi.org/10.1016/j.apergo.2021.103521>.

45. Perrot S, Lantéri-Minet M. Patients' Global Impression of Change in the management of peripheral neuropathic pain: clinical relevance and correlations in daily practice. *Eur J Pain*. 2019;23(6):1117–28. <https://doi.org/10.1002/ejp.1378>.
46. Petersen EA, Stauss TG, Scowcroft JA, Brooks ES, White JL, Sills SM, ... Mekhail NA. Effect of high-frequency (10-kHz) spinal cord stimulation in patients with painful diabetic neuropathy: a randomized clinical trial. *JAMA Neurol*. 2021;78(6):687–698. <https://doi.org/10.1001/jamaneurol.2021.0538>.
47. Rosser BA, Eccleston C. Smartphone applications for pain management. *J Telemed Telecare*. 2011;17(6):308–12. <https://doi.org/10.1258/jtt.2011.101102>.
48. Sayed D, Kallewaard JW, Rotte A, Jameson J, Caraway D. Pain relief and improvement in quality of life with 10 kHz SCS therapy: summary of clinical evidence. *CNS Neurosci Ther*. 2020;26(4):403–15. <https://doi.org/10.1111/cns.13285>.
49. Shealy CN, Mortimer JT, Reswick JB. Electrical inhibition of pain by stimulation of the dorsal columns: preliminary clinical report. *Anesth Analg*. 1967;46(4):489–91.
50. Simpson EL, Duenas A, Holmes MW, Papaioannou D, Chilcott J. Spinal cord stimulation for chronic pain of neuropathic or ischaemic origin: systematic review and economic evaluation. *Health Technol Assess*. 2009;13(17):iii, ix–x, 1–154. <https://doi.org/10.3310/hta13170>.
51. Taylor RS, Van Buyten JP, Buchser E. Spinal cord stimulation for chronic back and leg pain and failed back surgery syndrome: a systematic review and analysis of prognostic factors. *Spine (Phila Pa 1976)*. 2005;30(1):152–60. <https://doi.org/10.1097/01.brs.0000149199.68381.fe>.
52. Tieppo Francio V, Polston KF, Murphy MT, Hagedorn JM, Sayed D. Management of chronic and neuropathic pain with 10 kHz spinal cord stimulation technology: summary of findings from preclinical and clinical studies. *Biomedicines*. 2021;9(6). <https://doi.org/10.3390/biomedicines9060644>.
53. Yongjun Z, Tingjie Z, Xiaoqi Y, Zhiying F, Feng Q, Guangke X, ... Yanqing L. A survey of chronic pain in China. *Libyan J Med*. 2021;15(1):1730550. <https://doi.org/10.1080/19932820.2020.1730550>.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

