

Accepted Manuscript

Title: Detection of azaspiracids in mussels using electrochemical immunosensors for fast screening in monitoring programs

Authors: Sandra Leonardo, Jane Kilcoyne, Ingunn A. Samdal, Christopher O. Miles, Ciara K. O'Sullivan, Jorge Diogène, Mònica Campàs



PII: S0925-4005(18)30316-2
DOI: <https://doi.org/10.1016/j.snb.2018.02.046>
Reference: SNB 24141

To appear in: *Sensors and Actuators B*

Received date: 28-11-2017
Revised date: 22-1-2018
Accepted date: 5-2-2018

Please cite this article as: Sandra Leonardo, Jane Kilcoyne, Ingunn A. Samdal, Christopher O. Miles, Ciara K. O'Sullivan, Jorge Diogène, Mònica Campàs, Detection of azaspiracids in mussels using electrochemical immunosensors for fast screening in monitoring programs, *Sensors and Actuators B: Chemical* <https://doi.org/10.1016/j.snb.2018.02.046>

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

Detection of azaspiracids in mussels using electrochemical immunosensors for fast screening in monitoring programs

Sandra Leonardo^a, Jane Kilcoyne^b, Ingunn A. Samdal^c, Christopher O. Miles^{c,d},

Ciara K. O'Sullivan^{e,f}, Jorge Diogène^a, Mònica Campàs^{a*}

^aIRTA, Ctra. Poble Nou, km. 5.5, 43540 Sant Carles de la Ràpita, Spain

^bMarine Institute, Rinville, Oranmore, County Galway H91 R673, Ireland

^cNorwegian Veterinary Institute, P.O. Box 750 Sentrum, N-0106 Oslo, Norway

^dNational Research Council Canada, 1411 Oxford St., Halifax, Nova Scotia, B3H 3Z1, Canada

^eDepartament d'Enginyeria Química, Universitat Rovira i Virgili, Av. Països Catalans, 26, 43007 Tarragona, Spain


^fInstitució Catalana de Recerca i Estudis Avançats, Pg. Lluís Companys, 23, 08010 Barcelona, Spain

*Corresponding author: Mònica Campàs

Tel.: +34 902 789 449 (ext. 1842); Fax: +34 977 744 138

E-mail address: monica.campas@irta.cat

Graphical abstract



Highlights

- The first electrochemical immunosensors for azaspiracids detection are reported.
- Bioaffinity interactions are used for the controlled immobilisation of the antibody.
- The strong avidin–biotin interaction allows the reusability of the immunosensors.
- The immunosensors detect all AZAs from far below to far above the regulatory limit.
- Excellent correlations with LC-MS/MS are obtained in the analysis of mussels.

Abstract

Given the widespread occurrence of azaspiracids (AZAs) it is clearly necessary to advance in simple and low-cost methods for the rapid detection of these marine toxins in order to protect seafood consumers. To address this need, electrochemical immunosensors for the detection of AZAs based on a competitive direct immunoassay using peroxidase-labelled AZA as a tracer were developed. An anti-AZA polyclonal antibody was immobilised in a controlled and stable manner on protein G or avidin-coated electrodes. Experimental conditions were first optimised using colorimetric immunoassays on microtitre plates, providing intermediate products already applicable to the accurate detection of AZAs. Then, transfer of the protein G and avidin–biotin interaction-based immunoassays to 8-electrode arrays provided compact and miniaturised devices for the high-throughput detection of AZAs. The low amounts of immunoreagents required as well as the potential for reusability of the avidin–biotin interaction-based immunosensors represented significant economic savings as well as a contribution to sustainability. The electrochemical immunosensors enabled the quantification of all regulated AZAs below the regulatory limit, as well as a broad range of other toxic AZA analogues (from 63 ± 3 to 2841 ± 247 $\mu\text{g AZA-1 equiv./kg}$ for the protein G-based immunosensor and from 46 ± 2 to 3079 ± 358 $\mu\text{g AZA-1 equiv./kg}$ for the avidin–biotin interaction-based immunosensor). The good agreement between the results obtained by the immunosensors and LC-MS/MS in the analysis of naturally contaminated mussel samples demonstrated the easy implementation of electrochemical immunosensors for routine analysis of AZAs in food safety monitoring programs.

Keywords: azaspiracid; antibody; electrochemical immunosensor; protein G; biotin; mussel.

1. INTRODUCTION

Harmful algal blooms (HABs) and their associated marine toxins pose a serious threat to human health and are an economic concern for the shellfish industry. Among many different groups of marine toxins of microalgal origin are azaspiracids (AZAs). AZAs were first identified in 1998 from contaminated mussels from Ireland that had caused a poisoning outbreak in The Netherlands in the mid-1990s [1]. Since then, AZAs have been particularly problematic in Ireland, but they have also been reported in phytoplankton and/or shellfish around the world including the US [2], China [3], Japan [4], Chile [5, 6] and Argentina [7]. In Europe, AZAs have been found in several countries adjacent to the North Sea such as Norway, Denmark, UK and Sweden [8-10]. AZAs have also been found in the Atlantic coast of other European countries including France, Portugal and Spain [11-15], in the Atlantic coast of NW Africa [16], as well as in countries of the Mediterranean Sea [17].

The AZA group includes more than 40 analogues, which are either produced by phytoplankton of the genera *Azadinium* and *Amphidoma*, through biotransformation in shellfish, or as by-products resulting from storage or cooking of AZA-contaminated shellfish [18]. However, only AZA-1–3 are currently regulated by the European Commission, with 160 µg AZA-1 equivalents/kg being the maximum permitted level in shellfish meat [19].

Although current reference chemical methods are highly specific and sensitive, they require the use of sophisticated equipment and trained personnel, being expensive and relatively slow to perform. Thus, there is a need for easy-to-use, rapid, inexpensive and accurate devices for the detection of AZAs in shellfish in monitoring programs. Biosensors have the potential to address this need and, among them, electrochemical biosensors stand out for several reasons: their inherently high sensitivities, low cost, possibility for miniaturisation of electrodes and potentiostats, compatibility with microfluidics systems and automation and subsequent simplification of the protocols [20]. Likewise, the high specificity and selectivity of antibodies (Abs) positions immunosensors as highly attractive candidates for the rapid screening of marine toxins.

Nonetheless, the lack of commercially available anti-AZA Abs has hindered the development of immunosensors, as well as immunoassays, for the detection of AZAs. To date, only a monoclonal antibody (MAb) raised using AZA-1 [21] and a polyclonal antibody (PAb) raised using a synthetic fragment of AZA [22] have been reported. While the MAb was used in the development of a flow fluorimetry-based immunoassay [23], the anti-AZA PAb has been used in the development

of an enzyme-linked immunosorbent assay (ELISA) [24] and an electrochemical immunoassay using magnetic beads as immunorecognition supports [25].

Another crucial step in the development of immunoassays and immunosensors is the immobilisation of the Ab on a solid support. To achieve full functionality, the conformation of the Abs should not be altered and their binding sites should remain accessible after immobilisation. Moreover, in the development of electrochemical biosensors, the immobilisation procedure should ensure a close proximity between the label and the transducer in order to obtain an efficient electron transfer. Bioaffinity immobilisation, mainly based on the avidin–biotin interaction and the affinity of protein A/G for immunoglobulins (IgGs), provides an attractive method for the controlled and stable surface-tethering of antibodies. In our previous work [25], the anti-AZA PAb was immobilised on protein G-coated magnetic beads and a competitive step using peroxidase-labelled AZA (AZA–HRP) was performed. The use of magnetic beads facilitated performance of the assay in suspension, thus allowing rapid assay kinetics, but mass transfer limitations were observed when the immunocomplexes were subsequently anchored on the electrode surface to perform electrochemical detection. In the present work, the anti-AZA PAb is immobilised on the electrode surface by means of protein G, or, alternatively, through the avidin–biotin interaction, with the aim of overcoming the drawbacks associated with the packed distribution of magnetic beads on the electrode surface and allowing a controlled and homogenous immobilisation of the antibody directly on the transducer surface. The immobilisation of the Ab directly on the transducer surface provides more compact and automated devices, since all reactions are performed on the electrode array, with possible signal enhancement since the enzyme product is concentrated closer to the transducer surface.

Colorimetric immunoassays were first developed on microtitre plates and used for protocol optimisation and assay characterisation. Their suitability for AZAs screening and quantification in mussels was also demonstrated. Protein G and avidin–biotin interaction-based immunoassays were then transferred to electrode arrays to develop the corresponding competitive electrochemical immunosensors. With the aim of further improving the economic saving represented by the use of the immunosensors, their reusability was explored. The immunosensors were applied to the determination of AZAs in a mussel certified reference material (CRM) and in mussel samples obtained from the Irish monitoring program. To the best of our knowledge, this work describes the first immunosensors for AZAs detection and guarantees their implementation in routine monitoring programs.

2. MATERIALS AND METHODS

2.1. Reagents and solutions

Protein G from *Streptococcus* sp., avidin from egg white, sodium carbonate, sodium bicarbonate, potassium phosphate monobasic, potassium chloride, Tween–20, bovine serum albumin (BSA) and 3,3'-5,5'-tetramethylbenzidine (TMB) liquid substrate were purchased from Sigma–Aldrich (Madrid, Spain). Milli-Q water (Millipore, Bedford, USA) was used to prepare solutions.

Certified reference materials (CRMs) of AZA-1–3, Zero-Mus and AZA-Mus were obtained from the National Research Council of Canada (NRC, Halifax, NS, Canada). Reference materials (RMs) of AZA-4–10 were prepared as described by Kilcoyne et al. [26].

HRP (Type VI-A) from Sigma–Aldrich (Oslo, Norway) was conjugated to purified AZA-1 [27] using standard procedures for covalent linking of small molecules to enzymes [28]. Antiserum AgR367-11b (anti-AZA PAb) was obtained after 11 immunisations with two different haptens as described in Samdal et al. [24].

Biotin labelling of the anti-AZA PAb was performed with EZ-Link™ NHS–PEG₄–Biotin from Thermo Fisher (Barcelona, Spain) following the manufacturer's manual. Unreacted NHS–PEG₄–Biotin was removed by Zeba Spin Desalting Columns (7 kDa MWCO, 2 mL) from Thermo Fisher.

2.2. Equipment

Colorimetric measurements were performed with a Microplate Reader KC4 from BIO-TEK Instruments, Inc. (Winooski, VT, USA). Gen5 software was used to collect and evaluate data.

An array of eight screen-printed carbon electrodes (DRP-8X110) and a boxed connector (DRP-CAST8X) were provided by Dropsens S.L. (Oviedo, Spain). The array consists of 8 carbon working electrodes of 2.5 mm in diameter, each with its own carbon counter electrode and silver reference electrode. Amperometric measurements were performed with a PalmSens potentiostat connected to an 8-channel multiplexer (MUX8) (Houte, The Netherlands). Data were collected and evaluated with PalmSens PC software.

2.3. Raw and heat-treated mussel tissues

AZA-Mus CRM (NRC, Halifax, NS, Canada) prepared from naturally contaminated mussel (*Mytilus edulis*) tissues and AZA-contaminated raw mussel samples (*M. edulis*) from the routine

monitoring program in Ireland, were selected for analysis. Zero-Mus CRM (*M. edulis*) (NRC, Halifax, NS, Canada) was used to evaluate matrix effects.

Mussels were shucked and homogenised with an Ultra Turrax homogenizer. A three-step extraction with MeOH (10 mL) was performed on AZA-Mus and Zero-Mus homogenised tissues (1 g) according to Gerssen et al. [29], using a protocol that was intra-laboratory validated by García-Altarex et al. [30]. A vortex-mixer MS2 Minishaker (IKA Labortechnik, Staufen, Germany) and a centrifuge Jouan MR 23i (Thermo Fisher Scientific Inc., Waltham, MA, USA) were used. Crude extracts were filtered through polytetrafluoroethylene (PTFE) 0.2- μ m membrane syringe filters. Thus, AZA-Mus and Zero-Mus extracts at a matrix concentration of 100 mg/mL were obtained.

Extraction of the Irish AZA-contaminated raw mussel samples was performed by a two-step extraction with MeOH (25 mL) as follows. Homogenised tissue samples were weighed (2 g) into 50-mL centrifuge tubes and extracted by vortex mixing for 1 min with 9 mL of MeOH, centrifuged at 3950 *g* (5 min), and the supernatants decanted into 25-mL volumetric flasks. The pellets were further extracted using an Ultra Turrax for 1 min with an additional 9 mL of MeOH, centrifuged at 3950 *g* (5 min), and the supernatants decanted into the same 25-mL volumetric flasks, which were brought to volume with MeOH. A portion (10 mL) of each extract was transferred into sealed centrifuge tubes and placed for 10 min in a water bath heated to 90 °C to decarboxylate carboxylated AZAs [31, 32]. The raw and heat-treated samples were then passed through Whatman 0.2- μ m cellulose acetate filters into HPLC vials for analysis. Heated and raw mussel extracts were obtained at a matrix concentration of 80 mg/mL. All samples were stored at -20 °C until analysis.

2.4. Colorimetric immunoassays protocol

Colorimetric immunoassays were carried out on 96-well microtitre plates. Microtitre wells were incubated with 50 μ L of 10 μ g/mL protein G or 1 μ g/mL avidin in 0.1 M carbonate buffer, pH 9.6, for 1 h. The anti-AZA PAb or biotinylated PAb was then immobilised on the protein G or avidin-coated plates, respectively, by the addition of 50 μ L of the corresponding antibody dilution (from 1/10000 to 1/80000 for protocol optimisation and 1/40000 for the final competition assay) in PBS–Tween (0.1 M PBS, pH 7.2, with 0.05% v/v Tween–20) for 1 h. Blocking was then carried out using 100 μ L of PBS–Tween containing 2% w/v BSA for 1 h. Subsequently, the competition step was performed using 25 μ L of AZA-1 standard solutions (from 0.20 μ g/L to 100 μ g/L) or natural samples at different dilutions in PBS–Tween and 25 μ L of different dilutions of AZA–HRP (from

1/800 to 1/3200 for protocol optimisation and 1/1600 for the final competition assay) for 30 min. Finally, 100 μL of TMB liquid substrate was added, and 10 min later the absorbance was read at 620 nm. After each step, wells were rinsed three times with 100 μL PBS–Tween. During incubations, microtitre plates were placed on a plate shaker. All reactions were carried out at room temperature.

2.5. Electrochemical immunosensors protocol

The immunosensor assay protocols were essentially the same as the colorimetric immunoassays except for adjustments to the volumes for 8-electrode arrays and differences in the detection step. Volumes of 10 μL were applied to each working electrode (5 μL of standard or sample dilution plus 5 μL of AZA–HRP in the competition step), and the blocking step was performed by immersion of the electrode arrays in PBS–Tween containing 2% w/v BSA. To perform the electrochemical measurement, 10 μL of TMB was added to each electrode and incubated for 10 min and, finally, the TMB reduction current was measured by applying -0.2 V vs. Ag for 0.5 s. After each step, the electrode arrays were rinsed with PBS–Tween and dried. All reactions were carried out at room temperature.

For the regeneration of the electrochemical immunosensors, the electrode arrays were rinsed with PBS–Tween after the electrochemical measurement and immersed in glycine buffer, pH 2.7, for 30 min. The immunosensors were then rinsed again with PBS–Tween and stored at 4 $^{\circ}\text{C}$ until use.

2.6. LC-MS/MS analysis

For LC-MS/MS analysis of AZA analogues, a Waters Acquity UPLC coupled to a Xevo G2-S QToF monitoring in MS^e mode (m/z 100–1200) was used with leucine enkephalin as the reference compound. The cone voltage was 40 V, collision energy was 50 V, the cone and desolvation gas flows were set at 100 and 1000 L/h, respectively, and the source temperature was 120 $^{\circ}\text{C}$. Analytical separation was performed on an Acquity UPLC BEH C18 (50 \times 2.1 mm, 1.7 μm) column (Waters). Binary gradient elution was used, with phase A consisting of H₂O and phase B of CH₃CN (95%) in H₂O (both containing 2 mM ammonium formate and 50 mM formic acid). The gradient was from 30–90% B over 5 min at 0.3 mL/min, held for 0.5 min, and returned to the initial conditions and held for 1 min to equilibrate the system. The injection volume was 2 μL and the column and sample temperatures were 25 $^{\circ}\text{C}$ and 6 $^{\circ}\text{C}$, respectively. AZA-1–3 were quantified relative to CRMs while AZA-4–10 were quantified with RMs [26]. Matrix interferences were

assessed using single point matrix matched standards for AZA-1–6 and AZA-8, using a blank *M. edulis* tissue. An aliquot (2 g) of tissue was extracted as described above, this time making the solution up to a final volume of 20 mL. The matrix-matched standard was prepared by adding 1.25 mL of an AZA-1–6 and AZA-8 stock solution in MeOH to 1 mL of the filtered (Whatmann, 0.2 μ m, cellulose acetate filter) blank tissue extract. In parallel, a MeOH standard was prepared by adding 1.25 mL of the same AZA-1–6 and AZA-8 stock solution to 1 mL of MeOH.

2.7. Data analysis

Measurements were performed in triplicate for the colorimetric immunoassays and electrochemical immunosensors and in duplicate for LC-MS/MS analysis. The immunoassay calibration curves were fitted using a sigmoidal logistic four-parameter equation. Linear regression was used to evaluate the correlation between AZA-1 equivalent concentrations in naturally contaminated mussel samples determined with the colorimetric immunoassays or the electrochemical immunosensors and the values obtained from the LC-MS/MS analyses. To evaluate differences between approaches, data were first tested for normality. To compare values from two different groups, the paired *t*-test was used for normally distributed data sets, while Wilcoxon matched-pairs signed-ranks was used for non-normally distributed data. One-way ANOVA was performed to compare the values obtained in the analysis of the mussel samples by the colorimetric immunoassays, electrochemical immunosensors and LC-MS/MS analysis. Differences were considered statistically significant at the 0.05 level. SigmaStat 3.1 was used for statistical analysis.

3. RESULTS AND DISCUSSION

3.1. Colorimetric immunoassays

Microtitre plates were coated with protein G or avidin by passive adsorption, taking advantage of the hydrophobic interactions and the electrostatic forces generated between the negatively charged proteins and the positively charged microplates in alkaline conditions. The anti-AZA PAb was then immobilised on the coated plates by means of the affinity of the protein G to the Fc region of the antibody or through the strong avidin–biotin interaction following antibody biotinylation (Fig. 1). Both bioaffinity interactions provided the stable immobilisation of the PAb, while retaining its biological activity. However, while the use of protein G favoured the optimum orientation of the antibody to achieve optimal antigen binding without requiring any chemical modification, the avidin–biotin interaction required biotin labelling of the primary amines of the PAb, typically distributed on the exterior of the entire antibody, which did not ensure the correct orientation of the antibody.




Fig. 1. Schematic representation of the immunoassays and immunosensors configurations achieved by the anti-AZA PAb immobilisation based on **(a)** protein G or **(b)** avidin–biotin affinity interaction on **(a)** microtitre plate wells and **(b)** electrodes, as examples.

Checkerboard titrations and competitive immunoassays were performed to determine if the antibody immobilisation was successful and to establish the optimum PAb and AZA–HRP concentrations. The absorbance values obtained showed expected trends according to antibody and AZA–HRP tracer concentrations. Lower absorbance values were obtained when the antibody was immobilised through the avidin–biotin interaction. Increasing the amount of avidin on the plate did not improve the immobilisation yield of the biotinylated antibody. Nevertheless, the lower absorbance values achieved in this configuration are not unexpected taking into account the biotinylation yield, which may not reach 100%, the non-optimally oriented immobilisation of the PAb and/or the possibility of biotinylation and immobilisation of other molecules containing primary amines present in the antiserum.

Both immobilisation strategies provided very low non-specific adsorption values of the AZA–HRP. Calibration curves demonstrated competition of free AZA-1 with AZA–HRP for PAb binding. Calibration curves were background-corrected with respect to the controls with no AZA-HRP and fitted to the sigmoidal logistic four-parameter equation:

$$y = y_0 + \frac{a}{1 + \left(\frac{x}{x_0}\right)^b}$$

where a and y_0 are the asymptotic maximum and minimum values, respectively, x_0 is the x value at the inflection point and b is the slope at the inflection point. Greater sensitivities were achieved with decreasing concentrations of antibody and tracer. Consequently, 1/40000 PAb and 1/1600 AZA–HRP dilutions were selected as a compromise between low antibody/tracer loading and sufficiently high absorbance values. Figure 2 shows the calibration curves for the optimised protein G and avidin-biotin interaction-based immunoassays. In table 1, limits of detection (LODs), established as the 10% inhibition coefficient (IC_{10}) and working ranges (between IC_{20} and IC_{80}) are presented together with the equations and the corresponding R^2 values. Differences between the two approaches were not significant ($t=0.292$, $P=0.774$). In comparison with the competitive colorimetric immunoassay previously reported by our group [25], where magnetic beads were used as antibody immobilisation supports, it was possible to use lower antiserum and tracer concentrations, which could explain the lower LODs achieved in these approaches.




Fig. 2. Colorimetric calibration curves for AZA-1 obtained using the protein G-based immunoassay (black) and the avidin–biotin interaction-based immunoassay (white) in buffer (circle) and in 20 mg/mL mussel matrix with 20% MeOH (triangle).

Table 1. Analytical parameters derived from the sigmoidal logistic four-parameter fitting for the Protein G-based and avidin–biotin interaction-based colorimetric immunoassays and electrochemical immunosensors.

		LOD (IC ₁₀) (µg/L)	Working range (IC ₂₀ –IC ₈₀) (µg/L)	Equation	R ²
Colorimetric immunoassays	Protein G	0.28	0.70 – 22.74	$y = 2.052 + \frac{99.494}{1 + (\frac{x}{3.468})^{0.805}}$	0.999
	Avidin–biotin	0.14	0.39 – 22.47	$y = 6.776 + \frac{94.670}{1 + (\frac{x}{1.999})^{0.751}}$	1.000
Electrochemical immunosensors	Protein G	0.61	1.25 – 56.81	$y = -10.073 + \frac{132.192}{1 + (\frac{x}{5.413})^{0.520}}$	0.999
	Avidin–biotin	0.37	0.92 – 61.58	$y = -74.378 + \frac{208.612}{1 + (\frac{x}{32.208})^{0.295}}$	0.999

3.2. Study of matrix effects

Prior to the analysis of mussel samples, AZA-1 calibration curves using a blank certified reference mussel tissue matrix (CRM-Zero-Mus) were performed to evaluate matrix effects on the immunoassays. A matrix concentration of 20 mg/mL mussel matrix was chosen according to the

protocol for lipophilic toxins extraction in shellfish (100 mg of matrix in 1 mL of MeOH) and its subsequent dilution to 20% v/v MeOH, a percentage that had been previously demonstrated not to interfere with the assay [25]. No significant differences were observed between the calibration curves performed in buffer and in 20 mg/mL mussel matrix with 20% v/v MeOH, neither in the protein G-based immunoassay ($t=0.24$, $P=0.81$) nor in the immunoassay based on the avidin–biotin interaction ($t=0.32$, $P=0.77$) (Fig. 2). Consequently, considering a 20 mg/mL matrix loading, effective LODs of $14 \pm 1 \mu\text{g AZA-1 equiv./kg}$ and $7 \pm 2 \mu\text{g AZA-1 equiv./kg}$ were achieved for the immunoassays developed by means of protein G and avidin–biotin immobilisation, respectively. Both immunoapproaches provided a broad working range, from $35 \pm 5 \mu\text{g AZA-1 equiv./kg}$ to $1137 \pm 150 \mu\text{g AZA-1 equiv./kg}$ in the case of the protein G-based immunoassay and from $20 \pm 3 \mu\text{g AZA-1 equiv./kg}$ to 1124 ± 118 in the case of the immunoassay based on the avidin–biotin interaction. Considering the EC regulatory limit for AZAs of $160 \mu\text{g/kg}$, the immunoassays enabled the quantification of mussel samples from far below to far above the regulatory limit without requiring additional sample dilutions.

3.3. Electrochemical immunosensors

After protocol optimisation using the colorimetric immunoassays, both strategies were transferred to 8-electrode arrays to develop the corresponding electrochemical immunosensors (Fig. 1). Although reagent concentrations were the same as those selected in the colorimetric immunoassays, the use of lower volumes on the screen-printed electrodes required 5-fold lower amounts of protein G, avidin, anti-AZA PAb and tracer, which represents a substantial economic improvement. After background correction with respect to the controls with no AZA-HRP and fitting the calibration curves to the sigmoidal logistic four-parameter equation, Protein G and avidin–biotin interaction-based immunosensors provided similar analytical performances ($t=0.10$, $P=0.92$) (Table 1, Fig. 3). This work presents the first immunosensors for AZAs reported to date, providing user-friendly and compact tools that favour operation with low sample volumes and the performance of multiple measurements in a short time.

Considering a 5-fold sample dilution to 20 mg/mL mussel matrix, as in the colorimetric assays, effective working ranges between 63 ± 3 and $2841 \pm 247 \mu\text{g AZA-1 equiv./kg}$ for the protein G-based immunosensor and between 46 ± 2 and $3079 \pm 358 \mu\text{g AZA-1 equiv./kg}$ for the avidin–biotin interaction-based immunosensor were calculated, again providing a broad working range that included the EC permitted threshold value.




Fig. 3. Electrochemical calibration curves for AZA-1 obtained using the protein G-based immunosensor (black) and the avidin–biotin interaction-based immunosensor (white) in buffer.

3.4. Regeneration of the electrochemical immunosensors

Electrochemical immunosensors are commonly developed using disposable screen-printed electrodes due to their low fabrication cost and the possibility of mass production. However, the lack of commercially available anti-AZA antibodies and the time-consuming antibody production process are clear limiting factors in the development of immunosensors for AZAs detection. The feasibility to reuse immunosensors presents expedient advantages regarding these limitations. With this purpose in mind, the possibility to remove the AZAs and the AZA–HRP bound to the PAb, once the competition assay had been performed, while retaining the PAb immobilisation on the electrode surface and its functionality was evaluated.

Antibody-antigen interaction usually occurs at physiological pH and ionic strength, such as in PBS, and can be disrupted by simply raising or lowering the pH or altering the ionic state, ideally releasing the antibody or antigen without irreversibly denaturing or inactivating them. Thus, when immunosensors containing AZA–HRP were immersed in glycine buffer, pH 2.7, for 30 min, only background currents were observed after TMB incubation, indicative of complete AZA–HRP elution from the electrode. A subsequent incubation with AZA–HRP resulted in a response of ~25% in the protein G-based functionalised electrodes, and a response of ~100% in the avidin–biotin configuration. This 100% response demonstrates that the biotinylated antibody was retained on the avidin-coated electrode retaining its functionality, while the 25% response observed in the protein G approach suggests partial co-elution of the antibody during the elution

step in this specific configuration. Responses close to 100% following AZA-HRP incubation were maintained after the regeneration of the functionalised electrodes through the avidin–biotin interaction for 6 consecutive times on 3 different days (Fig. 4). These results demonstrate the reusability of the avidin–biotin interaction-based immunosensors and also highlight the storage stability of the immunosensors for at least 3 days. The possibility to reuse and store the immunosensors until use avoids the immobilisation of additional anti-AZA PAb amounts on the electrode surface, as well as simplifies and shortens the protocol assay.




Fig. 4. Electrochemical responses of the avidin–biotin interaction-based immunosensors achieved after surface regeneration for 6 consecutive measurements (M#1–M#6) on 3 different days.

3.5. AZAs detection in mussel samples

To demonstrate the applicability of the electrochemical immunosensors as well as the colorimetric immunoassays to the determination of AZAs in shellfish, naturally contaminated mussel (*M. edulis*) samples were analysed using the different approaches.

First, AZAs in a CRM containing AZA-1–3 (certified concentrations) and other analogues (AZA-4–10) (non-certified concentrations) were determined by the protein G and avidin–biotin interaction-based colorimetric immunoassays and electrochemical immunosensors (Table 2). The immunoapproaches provided a global quantification relative to AZA-1, which was compared with the sum of all different analogues concentrations provided by the NRC. Correlations of 153.4 % and 152.7 % were achieved between the quantifications provided by the protein G and avidin–biotin interaction-based colorimetric immunoassays, respectively, and NRC values.

Similar correlations were also obtained in the quantifications provided by the protein G (153.6 %) and avidin-biotin (154.8 %) electrochemical immunosensors. These percentages higher than 100% are not surprising taking into account the different recognition principles between analytical methods. While NRC values are obtained by physicochemical approaches, the immunoapproaches provide responses based on the structural recognition of the toxins by the antibody, which can differ between analogues and the assay configuration [25]. Consequently, the application of cross-reactivity factors (CRFs) to each individual AZA analogue concentration can contribute to better understand the correlation between the different analytical methods. Thus, CRFs that had been previously established in the magnetic bead-based immunoassay [25] –where the same PAb and a similar immunological approach were used– were applied to the individual AZA analogue concentrations provided by the NRC. Following the application of the corresponding CRFs to the certified values, correlations of 101.7% and 101.3% were achieved between the quantifications provided by the protein G and avidin–biotin interaction-based colorimetric immunoassays, respectively, and NRC values. Excellent correlations of 101.9 % and 102.7 % were also obtained in the quantifications provided by the protein G and avidin–biotin electrochemical immunosensors. The obtained results suggest that the PAb recognise all AZA analogues with the same cross-reactivity as in the magnetic bead-based immunoassay and highlight the potential application of all the developed immunoapproaches to the determination of AZAs in mussel samples.

In Ireland, AZAs have been detected in shellfish above the regulatory limit almost every year since the Irish monitoring program was established in 2001 [33]. Raw mussel samples (n=16) from this routine monitoring program containing a wide range of AZAs concentrations were selected for their quantification. From these 16 naturally contaminated samples, 11 had been previously analysed by the magnetic bead-based immunoassay [25]. Five additional mussel samples containing AZA-1, AZA-2 and AZA-3 levels below the established 160 µg/kg regulatory limit were included in the analysis to ensure that the developed immunoassays and immunosensors were able to provide reliable AZA quantifications at low concentrations (Table 1). No differences were observed between the quantifications provided by the protein G and avidin-biotin interaction-based colorimetric immunoassays and electrochemical immunosensors ($P=1.00$), regardless of whether the values were below ($P=0.851$) or above ($P=0.953$) the regulatory level. Standard deviations lower than 10% were obtained for all samples using both approaches.

Since the PAb used in the development of the immunoapproaches was able to recognise AZA carboxy cogeners in addition to other AZA analogues, the raw mussel samples analysed were

heated to perform LC-MS/MS analysis. Heating catalyses the decarboxylation of AZA carboxy congeners (e.g., AZA-17, AZA-19, AZA-21 and AZA-23) that may be present in the samples to AZA-3, AZA-6, AZA-4 and AZA-9 respectively [31, 32]. AZA quantifications provided by the protein G-based colorimetric immunoassay ($y = 1.489x + 42$, $R^2 = 0.981$) and electrochemical immunosensor ($y = 1.604x + 7$, $R^2 = 0.990$) correlated with those obtained by LC-MS/MS analysis ($p < 0.0001$). The immunoassay ($y = 1.574x + 18$, $R^2 = 0.990$) and the immunosensor ($y = 1.582x + 21$, $R^2 = 0.977$) developed through the avidin–biotin interaction also provided AZA quantifications in correlation with those achieved by the reference method ($p < 0.0001$) (Table 1, Figure 5). Without the application of the CRFs, the trend observed in the analysis of the mussel samples from the Irish monitoring program was similar to the trend obtained in the analysis of the CRM, being the quantifications achieved by the immunoapproaches ~1.5-fold those obtained with the reference method. Following the application of the corresponding CRFs to the individual contents determined by LC-MS/MS, AZA quantifications provided by the protein G-based immunoassay ($y = 0.965x + 14$, $R^2 = 0.999$) and immunosensor ($y = 1.035x - 19$, $R^2 = 0.999$) and the avidin–biotin interaction-based immunoassay ($y = 1.016x - 6$, $R^2 = 0.998$) and immunosensor ($y = 1.029x - 10$, $R^2 = 0.999$) were in excellent agreement with LC-MS/MS values (Table 2, Figure 5), with no significant differences observed between the quantifications achieved by any of the immunoapproaches and LC-MS/MS analysis ($P = 1.000$).

All these results indicate without doubt that the developed immunoassays and immunosensors are reliable tools for the screening and quantification of AZAs, facilitating not only the detection of all the regulated AZAs below the regulatory limit, but also other toxic analogues in a simple, rapid and cost-effective manner.

Table 2. AZA concentrations (μg AZA/kg mussel) of a mussel certified reference material (CRM) and 16 naturally contaminated mussel tissues from the routine monitoring program in Ireland by LC-MS/MS, the colorimetric immunoassays and the electrochemical immunosensors.

Samples	LC-MS/MS											Colorimetric immunoassays		Electrochemical immunosensors		
	AZA-1	AZA-2	AZA-3	AZA-4	AZA-5	AZA-6	AZA-7	AZA-8	AZA-9	AZA-10	Σ AZAs	AZA-1 equiv. (Σ AZAs applying CRFs)*	Protein G	Avidin-biotin	Protein G	Avidin-biotin
													AZA-1 equiv.	AZA-1 equiv.	AZA-1 equiv.	AZA-1 equiv.
CRM	1160	273	211	170	40	90	20	30	40	20	2054	3096	3150	3137	3155	3180
Monitoring																
S#1	24	5	15	nd	nd	3	nd	nd	nd	nd	47	77	77	82	79	66
S#2	22	8	15	nd	nd	3	nd	nd	nd	nd	48	77	79	73	80	76
S#3	28	8	17	nd	nd	2	nd	nd	nd	nd	55	86	94	89	82	86
S#4	21	8	23	nd	nd	3	nd	nd	nd	nd	55	98	95	97	113	65
S#5	33	10	19	nd	nd	2	nd	nd	nd	nd	64	98	105	103	104	100
S#6	49	19	31	nd	1	7	nd	2	nd	nd	109	172	174	177	146	171
S#7	55	19	29	1	1	4	nd	nd	nd	nd	109	165	171	124	159	165
S#8	53	20	31	1	2	4	nd	nd	nd	nd	111	170	133	158	171	130
S#9	128	45	66	3	4	10	nd	5	nd	nd	261	396	412	394	376	414
S#10	148	35	55	3	6	16	nd	1	1	3	268	399	381	391	398	404
S#11	144	59	80	13	5	12	nd	6	1	2	322	515	523	535	513	475
S#12	370	125	216	81	23	44	2	21	19	6	907	1623	1667	1534	1577	1646
S#13	425	139	189	77	17	35	4	21	16	4	927	1558	1561	1582	1550	1676
S#14	573	164	174	71	13	37	2	20	17	3	1074	1656	1631	1784	1661	1694
S#15	798	258	480	142	32	101	2	46	32	7	1898	3356	3202	3301	3399	3488
S#16	1524	414	309	132	33	64	10	49	38	9	2582	3639	3509	3778	3872	3664

* Applied CRFs are: AZA-1 = 1; AZA-2 = 0.76; AZA-3 = 2.73; AZA-4 = 3.83; AZA-5 = 1.39; AZA-6 = 2.70; AZA-7 = 2.00; AZA-8 = 1.85; AZA-9 = 2.69; AZA-10 = 2.17 [25].




Figure 5. Linear regressions for the correlations between sample quantifications by the colorimetric immunoassays or electrochemical immunosensors and the sum of AZA-1-10 analogues determined by LC-MS/MS analysis, before (white) and after (black) the application of the cross-reactivity factors (CRFs). Dashed lines represent the prediction intervals of 95 %.

4. CONCLUSIONS

The immobilisation of the anti-AZA PAb on protein G or avidin-coated supports via bioaffinity interactions has been used to develop immunoassays and immunosensors with excellent analytical performance, thanks to a controlled and stable antibody immobilisation. All approaches showed a broad working range that enabled the quantification of the current regulated AZAs below the regulatory threshold, but also a broad range of other toxic AZA analogues. No matrix effects were observed and no evaporation of the sample extracts was required, resulting in accurate quantifications with simple and rapid protocols. Although both colorimetric immunoassays and electrochemical immunosensors have been demonstrated to be powerful analytical methods for the reliable determination of AZAs, electrochemical immunosensors provide compact and miniaturised devices that pave the way towards the development of portable tools for *in situ* measurements. Moreover, the possibility to use lower amounts of immunoreagents, as well as the feasibility to reuse the avidin-biotin interaction-based immunosensors provide clear advantages in terms of sustainability and cost-effectiveness.

The good results obtained in the analysis of a considerable amount of naturally contaminated mussel samples evidence that electrochemical immunosensors for the detection of AZAs can be effectively implemented as screening tools in routine monitoring programs, as they provide easy-to-handle, rapid and low-cost high-throughput systems for the specific detection of AZAs in complex matrices. For their practical application in current monitoring programs, it is important to keep in mind that the regulatory limit of 160 µg AZA-1 equivalents/kg shellfish established by the European Commission only considers AZA-1–3 (including their toxic potential). As a consequence, in samples where more than these three analogues are present, immunoapproaches could provide results that may lead to a closure of a shellfish harvesting area, while LC-MS/MS analysis could not. This fact is not detrimental for immunoapproaches, but helps to better protect consumer health. Taking this in mind, we propose to establish a “positive”, “negative” but also a “suspicious” range to classify samples according to their AZA content. In the case of a “positive” result, a preventive closure of the shellfish harvesting area will be recommended to protect the consumer health and the sample will be analysed by LC-MS/MS to confirm the result. In the case of a “suspicious” sample, analysis by LC-MS/MS will help to determine the decision to undertake. By combining screening and confirmatory methods a faster and cost-effective system for marine toxin control in shellfish will be clearly achieved.

Acknowledgments

The research leading to these results has received funding under the *Sea Change* strategy with the support of the Marine Institute and the Marine Research Sub-Programme of the National Development Plan 2007–2013, co-financed under the European Regional Development Fund (ASTOX 2, (PBA/AF/08/001(01))). The publication reflects the views only of the author, and the European Union cannot be held responsible for any use that may be made of the information contained therein. The authors acknowledge financial support from CERCA Programme / Generalitat de Catalunya and Laura Ibañez for her help in the experimental work. Sandra Leonardo acknowledges scholarship from IRTA-Universitat Rovira i Virgili-Banco Santander (2013PIPF URV-IRTA-BS-01).

References

- [1] T. McMahon, J. Silke, Winter toxicity of unknown aetiology in mussels, *Harmful Algae News*, 14(1996).
- [2] V.L. Trainer, L. Moore, B.D. Bill, N.G. Adams, N. Harrington, J. Borchert, et al., Diarrhetic shellfish toxins and other lipophilic toxins of human health concern in Washington State, *Marine Drugs*, 11(2013) 1815-35.
- [3] J. Yao, Z. Tan, D. Zhou, M. Guo, L. Xing, S. Yang, Determination of azaspiracid-1 in shellfishes by liquid chromatography with tandem mass spectrometry, *Chinese journal of chromatography*, 28(2010) 363-7.
- [4] R. Ueoka, A. Ito, M. Izumikawa, S. Maeda, M. Takagi, K. Shin-Ya, et al., Isolation of azaspiracid-2 from a marine sponge *Echinoclathria* sp as a potent cytotoxin, *Toxicon*, 53(2009) 680-4.
- [5] G. Alvarez, E. Uribe, P. Avalos, C. Marino, J. Blanco, First identification of azaspiracid and spirolides in *Mesodesma donacium* and *Mulinia edulis* from Northern Chile, *Toxicon*, 55(2010) 638-41.
- [6] A. Lopez-Rivera, K. O'Callaghan, M. Moriarty, D. O'Driscoll, B. Hamilton, M. Lehane, et al., First evidence of azaspiracids (AZAs): A family of lipophilic polyether marine toxins in scallops (*Argopecten purpuratus*) and mussels (*Mytilus chilensis*) collected in two regions of Chile, *Toxicon*, 55(2010) 692-701.
- [7] A.D. Turner, A.B. Goya, Occurrence and profiles of lipophilic toxins in shellfish harvested from Argentina, *Toxicon*, 102(2015) 32-42.
- [8] K.J. James, A. Furey, M. Lehane, H. Ramstad, T. Aune, P. Hovgaard, et al., First evidence of an extensive northern European distribution of azaspiracid poisoning (AZP) toxins in shellfish, *Toxicon*, 40(2002) 909-15.
- [9] T. Torgensen, N.B. Bremnes, T. Rundberget, T. Aune, Structural confirmation and occurrence of azaspiracids in Scandinavian brown crabs (*Cancer pagurus*), *Toxicon*, 51(2008) 93-101.
- [10] B. Krock, U. Tillmann, T.J. Alpermann, D. Voss, O. Zielinski, A.D. Cembella, Phycotoxin composition and distribution in plankton fractions from the German Bight and western Danish coast, *Journal of Plankton Research*, 35(2013) 1093-108.
- [11] Z. Amzil, M. Sibat, F. Royer, V. Savar, First report on azaspiracid and yessotoxin groups detection in French shellfish, *Toxicon*, 52(2008) 39-48.
- [12] P. Vale, R. Bire, P. Hess, Confirmation by LC-MS/MS of azaspiracids in shellfish from the Portuguese north-western coast, *Toxicon*, 51(2008) 1449-56.

- [13] A.B. Magdalena, M. Lehane, S. Kryszewski, M.L. Fernandez, A. Furey, K.J. James, The first identification of azaspiracids in shellfish from France and Spain, *Toxicon*, 42(2003) 105-8.
- [14] U. Tillmann, D. Jaen, L. Fernandez, M. Gottschling, M. Witt, J. Blanco, et al., *Amphidoma languida* (Amphidomatacea, Dinophyceae) with a novel azaspiracid toxin profile identified as the cause of molluscan contamination at the Atlantic coast of southern Spain, *Harmful Algae*, 62(2017) 113-26.
- [15] J. Blanco, F. Arevalo, A. Morono, J. Correa, S. Muniz, C. Marino, et al., Presence of azaspiracids in bivalve molluscs from Northern Spain, *Toxicon*, 137(2017) 135-43.
- [16] H. Taleb, P. Vale, R. Amanhir, A. Benhadouch, R. Sagou, A. Chafik, First detection of azaspiracids in mussels in north west Africa, *Journal of Shellfish Research*, 25(2006) 1067-70.
- [17] R. Rossi, C. Dell'Aversano, B. Krock, P. Ciminiello, I. Percopo, U. Tillmann, et al., Mediterranean *Azadinium dexteroporum* (Dinophyceae) produces six novel azaspiracids and azaspiracid-35: a structural study by a multiplatform mass spectrometry approach, *Analytical and Bioanalytical Chemistry*, 409(2017) 1121-34.
- [18] P. Hess, P. McCarron, B. Krock, J. Kilcoyne, C.O. Miles, Azaspiracids: chemistry, biosynthesis, metabolism, and detection, in: L.M. Botana (Ed.), *Seafood and Freshwater Toxins. Pharmacology, Physiology, and Detection*. CRC Press, Boca Raton, FL, USA, 2014, pp. 799-822.
- [19] European Commission Regulation (EC) No 853/2004 of 29 April 2004 laying down specific hygiene rules for food of animal origin, *Official Journal of the European Union* L139, 55-205.
- [20] S. Leonardo, A. Toldrà, M. Campàs, Trends and prospects on electrochemical biosensors for the detection of marine toxins, in: J. Diogène, M. Campas (Eds.), *Recent Advances in the Analysis of Marine Toxins*, *Comprehensive Analytical Chemistry*, 78, 2017, pp. 303-341 .
- [21] M.O. Frederick, S.D.L. Marin, K.D. Janda, K.C. Nicolaou, T.J. Dickerson, Monoclonal antibodies with orthogonal azaspiracid epitopes, *Chembiochem*, 10(2009) 1625-9.
- [22] C.J. Forsyth, J. Xu, S.T. Nguyen, I.A. Samdal, L.R. Briggs, T. Rundberget, et al., Antibodies with broad specificity to azaspiracids by use of synthetic haptens, *Journal of the American Chemical Society*, 128(2006) 15114-6.
- [23] L.P. Rodriguez, N. Vilarino, M. Carmen Louzao, T.J. Dickerson, K.C. Nicolaou, M.O. Frederick, et al., Microsphere-based immunoassay for the detection of azaspiracids, *Analytical Biochemistry*, 447(2014) 58-63.
- [24] I.A. Samdal, K.E. Lovberg, L.R. Briggs, J. Kilcoyne, J. Xu, C.J. Forsyth, et al., Development of an ELISA for the Detection of Azaspiracids, *Journal of Agricultural and Food Chemistry*, 63(2015) 7855-61.
- [25] S. Leonardo, M. Rambla-Alegre, I.A. Samdal, C.O. Miles, J. Kilcoyne, J. Diogène, et al., Immunorecognition magnetic supports for the development of an electrochemical immunoassay for azaspiracid detection in mussels, *Biosensors & Bioelectronics*, 92(2017) 200-6.
- [26] J. Kilcoyne, M.J. Twiner, P. McCarron, S. Crain, S.D. Giddings, B. Foley, et al., Structure elucidation, relative LC-MS response and in vitro toxicity of azaspiracids 7-10 isolated from mussels (*Mytilus edulis*), *Journal of Agricultural and Food Chemistry*, 63(2015) 5083-91.
- [27] J. Kilcoyne, A. Keogh, G. Clancy, P. LeBlanc, I. Burton, M.A. Quilliam, et al., Improved isolation procedure for azaspiracids from shellfish, structural elucidation of azaspiracid-6, and stability studies, *Journal of Agricultural and Food Chemistry*, 60(2012) 2447-55.
- [28] G.T. Hermanson, *Bioconjugate Techniques*, Academic Press, London, 2008.
- [29] A. Gerssen, P.P.J. Mulder, M.A. McElhinney, J. de Boer, Liquid chromatography-tandem mass spectrometry method for the detection of marine lipophilic toxins under alkaline conditions, *Journal of Chromatography A*, 1216(2009) 1421-30.
- [30] M. Garcia-Altare, J. Diogene, P. de la Iglesia, The implementation of liquid chromatography tandem mass spectrometry for the official control of lipophilic toxins in seafood: Single-laboratory validation under four chromatographic conditions, *Journal of Chromatography A*, 1275(2013) 48-60.

[31] P. McCarron, J. Kilcoyne, C.O. Miles, P. Hess, Formation of azaspiracids-3,-4,-6, and-9 via decarboxylation of carboxyazaspiracid metabolites from shellfish, *Journal of Agricultural and Food Chemistry*, 57(2009) 160-9.

[32] J. Kilcoyne, P. McCarron, P. Hess, C.O. Miles, Effects of heating on proportions of azaspiracids 1-10 in mussels (*Mytilus edulis*) and identification of carboxylated precursors for azaspiracids 5, 10, 13, and 15, *Journal of Agricultural and Food Chemistry*, 63(2015) 10980-7.

[33] Marine Institute Ireland, Latest shellfish safety data. <http://www.marine.ie/Home/site-area/data-services/interactive-maps/latest-shellfish-safety-data/>, 2017 (accessed 17.11.17).

Author Biographies

Sandra Leonardo is a PhD student at IRTA in collaboration with the Department of Chemical Engineering of Universitat Rovirari Virgili in Tarragona. Her thesis focuses on the development of colorimetric assays and electrochemical biosensors for marine environmental monitoring and seafood safety. She obtained her degree in Biomedical Science in 2012 and acquired a master degree in Genetics and Genomics in 2013 from the University of Barcelona.

Jane Kilcoyne received her PhD in 2015 from Dublin Institute of Technology, Ireland. She currently works at the Marine Institute in Ireland, performing research on algal toxins and coordinating the research project MARBioFEED (enhanced biorefining for the production of marine biotoxins and microalgal fish feed). Her work includes the analysis of shellfish and algal extracts, culturing of microalgae, purification and structural elucidation of algal toxins.

Dr. **Ingunn Anita Samdal** is a researcher in the Toxinology Research Group at the Norwegian Veterinary Institute in Oslo. She obtained a Cand. Scient. in Biochemistry at the University of Oslo in 1999 and a Dr. Scient.-degree at the Norwegian School of Veterinary Science in Oslo in 2005. Her research focuses on secondary metabolites of marine and freshwater algae – their isolation, purification and detection. This with a special interest and expertise in developing antibodies to marine and freshwater biotoxins and using them for immunological detection and analysis methods such as ELISAs. She has developed specific antibodies to pinnatoxins, azaspiracids, microcystins, okadaic acid and yessotoxins, and used this to develop ELISAs.

Chris Miles is a Research Officer at the National Research Council Canada in Halifax, and was formerly a researcher at the National Veterinary Institute in Oslo. He received a PhD in chemistry from the University of Waikato, Hamilton, New Zealand in 1987. His work currently focusses on the chemistry of marine and freshwater algal toxins and production of certified reference materials.

Dr. **Ciara K. O' Sullivan** is a research professor from the Catalan Institute for Advanced Research Studies (ICREA). She is located at the Universitat Rovira i Virgili in Tarragona, where she heads the consolidated research group, INTERFIBIO, which has been awarded a certificate of excellence by the Catalan government. She has published >180 peer reviewed articles and her h-index is 40.

Jorge Diogène (Barcelona, 1962) is currently Senior Research Scientist and Director of the Marine Monitoring Subprogram at IRTA (Catalonia, Spain) dealing with the food safety of aquaculture and fishery products in Catalonia. Jorge Diogène has a PhD. in Applied and Fundamental Toxicology (Université Paris 7, France, 1993), Master in Marine Biology (University of South Alabama, USA, 1988) and graduate studies in Biology (Universitat de Barcelona, Spain, 1985). His research interests are oriented towards the evaluation of toxin production by microalgae and their presence in fish, shellfish and the environment, and method development for marine toxin quantification and risk assessment, including cell-based assays used to identify and quantify the toxicological potential of toxins. His research has contributed to a better understanding of several groups of toxins (Ciguatera, PSP, DSP, Cyclic Imines) and toxin-producing microalgae in different geographical areas (Mediterranean, North Atlantic, Macaronesia, Caribbean, Indian Ocean, Chile).

Dr. **Mònica Campàs** is researcher at IRTA, where she leads a research line on biosensors for the detection of toxins, toxic phytoplankton and other analytes. Mònica obtained a BSc in Chemistry in 1996 and a PhD in Chemical Engineering in 2002 from the Universitat Rovira i Virgili. Her research interests lie in the development of colorimetric assays and electrochemical biosensors using new biorecognition elements, immobilisation techniques and supports, transduction schemes and signal amplification strategies. Presently, her work focuses on the development and applicability of bioanalytical systems for the detection of emerging marine toxins (e.g. tetrodotoxins), toxic microalgae (e.g. *Karlodinium*, *Ostreopsis* and *Gambierdiscus*) and virus (e.g. oyster herpes virus). She has published 59 articles, 10 book chapters and 1 book. She has been the principal investigator of 10 projects. She has directed 2 PhD theses, 2 more being in progress.